



Client Services Manual

May 2025 Edition

Utah Public Health Laboratory Environmental Chemistry



The Environmental Chemistry Laboratory is TNI-accredited. TNI (The NELAC Institute) promotes the generation of data of known and documented quality through the National Environmental Laboratory Accreditation Program, or NELAP, and is considered the gold standard for environmental testing laboratories.

Quick Links

[Basic Information](#)

[Use Of This Client Services Manual](#)

[General Lab Practices & Policies](#)

[Sampling Request Instructions](#)

[Test Request Forms](#)

[Chain Of Custody \(COC\)](#)

Analytical Services

[Inorganic Chemistry](#)

[Metals](#)

[Organic Chemistry](#)

[Environmental Water Microbiology](#)

[Indexes](#)


BASIC INFORMATION

Address: Utah Public Health Lab
4431 S 2700 W
Taylorsville, UT 84129
Phone: (801) 965-2400
Website: <https://uphl.utah.gov/>

Laboratory business hours:
Mon. – Fri., 8:00 am – 5:00 pm

After hours, holidays, and emergency:
1-888-374-8824

To schedule sample testing:
801-965-2400 or 801-965-2405



A listing of all Environmental Chemistry testing, including methods and fees, is located on the UPHL website at: <https://uphl.utah.gov/uphl-service-fee-schedule/>.

USE OF THIS CLIENT SERVICES MANUAL

This manual is designed to provide a resource for sample collection and available testing information. Its contents are not to be used for regulatory purposes other than providing proper sample collection and preservation information.

To find a specific test, refer to the Analytical Services Section or the General Index toward the end of the manual. The Analytical Services Section is organized by testing group (Inorganic, Metals, Organic, Environmental Water Micro). Within each group, contents are organized alphabetically by contaminant. The index located at the end of the manual lists each test alphabetically along with their method numbers. Program specific indices are organized by the specific program, giving the analytes and tests codes.

Once a contaminant has been selected, refer to the sampling/testing instructions located on the page dedicated to the test.

GENERAL LAB PRACTICES & POLICIES

Our laboratory is responsible for the receipt of samples and testing of chemicals and environmental microbial contaminants in drinking water, wastewater, environmental soils and hazardous waste. Testing is done to ensure compliance with health and safety standards established by Federal and Utah State agencies. Services are provided to Utah DEQ, public water and wastewater utilities, local health departments, as well as other state and federal agencies.

Labeled sample collection materials such as water bottles, glass vials, and solids containers may be obtained from the laboratory. The laboratory tests each lot of containers to ensure they are free of contamination. The tested lot number appears on each container label.

Sample submission forms are available from the UPHL website at https://uphl.utah.gov/environmental-chemistry-program/Env_Chem_TestRequestForms/ or provided when containers are picked up. Follow sampling instructions as stated. The completed sample submission forms should accompany the samples when dropped off at the lab. All information should be checked and confirmed to be accurate before samples are submitted for testing. Tests, which are needed by the client but are not performed in our laboratory, will be subcontracted to a commercial laboratory with approval from the client. It should be noted that subcontracted testing will be invoiced at the subcontractor pricing.

Test results are provided to the submitting client, to other individuals as authorized by the submitting client, and to state and/or federal regulatory agencies as required by law. Fees for laboratory services will be charged to the submitting client.

SAMPLING REQUEST INSTRUCTIONS

To help reduce delays in processing your samples, please fill out one test form for each sample site. One sample site may have multiple bottles, but one form cannot have multiple collection sites. Complete the collection information on all of the bottles that are submitted and place completed form in a resalable bag. The Division of Drinking Water has requested some new fields on the forms—Facility ID and Sampling Point ID. If you are not certain what to put in those fields, please contact The Division of Drinking Water. Also, when making a sample delivery, please include a telephone number that is reachable 24-28 hours after sample has been received.

Please make sure you are using our current shipping address to prevent extra courier charges and delays in your shipment as well as to ensure that your environmental samples get delivered to the proper location once they are received at the lab, use one of these two

addresses listed below. If the samples need to go to someone's attention at the lab, please contact them prior to shipping so they can coordinate with sample receiving.

Address options for shipping environmental samples:

FedEx or UPS	USPS
Utah Public Health Laboratory Attn BCES SampleReceiving 4431 S 2700 W Taylorsville, UT84129	Utah Public Health Laboratory Attn BCES SampleReceiving PO Box 144300 Salt Lake City, UT 84131-9988
Utah Public Health Laboratory 4431 S 2700W Rm 149 Taylorsville, UT84129	Utah Public Health Laboratory Rm 149 PO Box 144300 Salt Lake City, UT 84131-9988

For questions, please contact:

Environmental Chemistry Sample Receiving
Utah Public Health Laboratory
4431 S 2700W
Taylorsville, UT84129
801-965-2405

Business hours are Monday-Friday from 8:00 am to 5:00 pm.

TEST REQUEST FORMS

When filling out test request forms the following fields of information must be completed for proper identification of samples at time of receipt:

- **System Name / Agency Name**
- **System Number / Agency Code**
- **Cost Code / Project Code**
- **Contact Information** (phone number is required)
- **Person Submitting Samples** (point of contact for clarifications)
- **Billing Information** (if submitting samples for the first time or if updates are needed)

If you are unsure of your system name and number or agency code, please contact the laboratory prior to sample submission. Or, for drinking water samples contact the Division of Drinking Water at 801-536-4200. If you do not have a system number and have not submitted samples to the laboratory before, a system number will be assigned at time of sample receipt. If you do not know

which cost code or project code applies to your samples, please contact the laboratory.

CHAIN OF CUSTODY

When submitting chain of custody samples, please complete the following steps to ensure proper preservation of sample integrity:

- **Place Seals on Sample Container Lid or Cap or over the cooler.** (must be initialed and dated at time of collection)
- **Identify Continuous Sample Possession** (signatures for dispatch, courier, relinquish, and so forth) located at bottom of chain-of-custody form
- **Verify Laboratory Receipt** (obtain copy of form when signed by DLS staff at time of receipt)

Note: An additional Chain of Custody request fee will be applied to the test request.

ANALYTICAL SERVICES

Inorganic Chemistry

Alkalinity	12
Ammonia	13
Biological Oxygen Demand (BOD, CBOD) - Subcontracted	14
Bromate	15
Bromide	16
Chlorate	15
Chloride	16
Chlorite	15
Color	17
Conductivity	18
Corrosivity	19
Cyanide	20
Fluoride	21
Nitrate + Nitrite	22
Nitrite	22
Nitrogen, Total	23
Odor	24
Perchlorate - Subcontracted	25
pH	26
Phosphorus, Total (All Forms)	27
Ortho-phosphate	XX
Silica	28
Sulfate	29
Sulfide	30
Solids - Dissolved (TDS), Suspended (TSS), Volatile (VSS)	31
Turbidity	32
Inorganic Testing Panels & Analytical Groupings	
Annual Inorganics and Metals (18 Parameters)	33
New Drinking Water Source (46 Parameters)	34
Chemistry Types (Analytical Groupings)	35
Nutrients, Total and Dissolved (Analytical Groupings)	37

[Metals](#)

[Organic Chem](#)

[Env Water Micro](#)

[Indexes](#)

ANALYTICAL SERVICES

Metals

Analytical Method for Metals	38
Digestion Methods for Metals (non-drinking water)	38
Metals	45
Aluminum (Al)	Manganese (Mn)
Antimony (Sb)	Mercury (Hg)
Arsenic (As)	39 Molybdenum (Mo)
Barium (Ba)	Nickel (Ni)
Beryllium (Be)	Potassium (K)
Cadmium (Cd)	Selenium (Se)
Calcium (Ca)	Silver (Ag)
Chromium (Cr)	Sodium (Na)
Cobalt (Co)	Strontium (Sr)
Copper (Cu)	Thallium (Th)
Iron (Fe)	Tin (Sn)
Lead (Pb)	Vanadium (V)
Magnesium (Mg)	Zirconium (Zr)
Chromium-VI - Subcontracted	41
Lead & Copper	42
Hardness	43
Metals Testing Panels & Analytical Groupings	
Annual Inorganics and Metals (18 Parameters)	34
New Drinking Water Source (46 Parameters)	35
Toxic Characteristic Leaching Procedure (TCLP-Metals)	44
Chemistry Types (Analytical Groupings)	36
Metals, Total and Dissolved (Analytical Groupings)	45

[Inorganic Chem](#)[Organic Chem](#)[Env Water Micro](#)[Indexes](#)

ANALYTICAL SERVICES

Organic
Chemistry

Chlorophyll-A (Pheophytin-A)	46
Cyanotoxins (Anatoxin-a, Cylindrospermopsin, Microcystins) by ELISA	47
Haloacetic Acids (HAAs) - Subcontracted	48
Organic Carbon , Total (TOC) & Dissolved (DOC)	49
Organic Constituents, UV-Absorbing (UV) - Subcontracted	50
Periphyton	51
Trihalomethanes (THMs) & Maximum Potential THMs - Subcontracted	52
BTEXN - Subcontracted	53
Carbamates - Subcontracted	54
Geosmin and MIB - Subcontracted	55
Herbicides - Subcontracted	56
Oil and Grease - Subcontracted	57
Pesticides and SVOCs - Subcontracted	58
Polychlorinated Biphenyls (PCBs) & Organochlorine Pesticides - Subcontracted	59
Surfactants - Subcontracted	60
Total Petroleum Hydrocarbons (TPH) - Subcontracted	61
Toxic Characteristic Leaching Procedure (TCLP-Organics) - Subcontracted	62
Volatile Organic Compounds (VOCs) - Subcontracted	63

Inorganic Chem

Metals

Env Water Micro

Indexes

ANALYTICAL SERVICES

Environmental Water Microbiology

Heterotrophic Plate Count (HPC)	64
Legionella	65
Total Coliforms/E.coli.	66

[Inorganic Chem](#)

[Metals](#)

[Organic Chem](#)

[Indexes](#)

INDEXES

Inorganic Chem	Program Specific: Clean Water Act (CWA)	67
Metals	Program Specific: Resource Conservation & Recovery Act (RCRA)	72
Organic Chem	Program Specific: Safe Drinking Water Act (SDWA)	77
Env Water Micro	General	82

Inorganic Chemistry

Name: Alkalinity
Method: SM 2320B

Application: Drinking Water, Surface Water



Analytes: Alkalinity
Carbonate Solids
Carbonate
Bicarbonate (BICD)
Carbon Dioxide
Hydroxide

Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: 120 mL Alkalinity – unpreserved plastic bottle
1L Total Chemistry – unpreserved plastic bottle (also applicable)

Preservative: Refrigerate or store on ice. Do not allow to freeze.

Handling: Recommend next day receipt at lab. Sample must be analyzed within 14 days of collection.

Method Technology: pH Titration and Calculation

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Ammonia [NH_3^+]

Method: SM4500-NH3 H

Groupings: Total Nutrients

Application: Drinking Water, Surface Water

Analytes: Ammonia



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container, be sure not to over fill to prevent loss of preservative.

Required Containers/Volume: 500 mL Nutrient Analysis – plastic bottle

Preservative: H_2SO_4 to pH <2, refrigerate or store on ice. Do not allow to freeze.

Handling: Recommend next day receipt at lab. Sample must be analyzed within 28 days after collection.

Method Technology: Distillation technique followed by Flow Injection Colorimetry

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Biological Oxygen Demand (BOD)
Method: SM 5210B

NOTE: This test is NOT performed at UPHL and will be sent to a subcontracted laboratory for testing.

Application: Drinking Water, Surface Water, Waste Water

Analytes: BOD5 (Biological Oxygen Demand)
CBOD (Carbonaceous BOD)
SBOD (Soluble BOD)
SCBOD (Soluble Carbonaceous BOD)

Note: Please take into account that the completion of the 5-day testing duration cannot fall on a weekend or holiday.



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: 2 L BOD – unpreserved plastic bottle

Preservative: Refrigerate or store on ice. Do not allow to freeze

Handling: Recommend same day receipt at lab. Sample testing must begin within 48 hours after collection and will be completed 5 days after testing begins.

Method Technology: Dissolved Oxygen Reduction over 5 days at 20°C

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Bromate [BrO_3^-],
Bromide [Br^-], Chlorate
[ClO_3^-], Chlorite [ClO_2^-]

Method: EPA 300.1



Groupings: Inorganic Disinfection By-Products

Application: Drinking Water, Surface Water

Analytes: Bromate
Bromide
Chlorate
Chlorite

Note: Prior to collection, analysis should be scheduled with the laboratory at (801) 965-2400

Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: 100 mL Bromate, Chlorate, Chlorite – plastic bottle

Preservative: 25 mg Ethylenediamine, refrigerate or store on ice. Do not allow to freeze.

Handling: **Recommend next day receipt at lab.** Sample testing must begin within 28 days after collection, except for Chlorite and Bromate which must begin within 14 days.

Method Technology: Ion Chromatography

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Bromide [Br⁻], Chloride [Cl⁻]

Methods: EPA 300.1, EPA 300.0, or EPA 325.2

Groupings: Disinfectants

Application: Drinking Water (EPA 300.1, EPA 300.0)
Surface Water (EPA 325.2)



Analytes: Bromide
Chloride

Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: 120 mL Bromide & Chloride – unpreserved plastic bottle
120 mL Sulfate – unpreserved plastic bottle (also applicable)
1 L Total Chemistry – unpreserved plastic bottle (also applicable)

Preservative: Refrigerate or store on ice. Do not allow to freeze.

Handling: **Recommend next day receipt at lab.** Sample testing must begin within 28 days after sample collection.

Method Technology: Ion Chromatography and Flow Injection Colorimetry

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Color
Method: SM 2120B

Groupings: New Drinking Water Source

Application: Drinking Water, Surface Water

Analytes: Color



Note: Prior to collection of Color, analysis must be scheduled at (801) 965-2400.

Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: 250 mL Color – plastic bottle

Preservative: Refrigerate or store on ice. Do not allow to freeze.

Handling: Recommend same day receipt at lab. Sample testing must begin within 48 hours after collection.

Method Technology: Flow Injection Colorimetry

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Conductivity
Method: SM 2510B

Groupings: Total Chemistry

Application: Total & Filtered Metals
Corrosivity

Analytes: L-Specific Conductance



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: 120 mL Conductivity – unpreserved plastic bottle
1 L Total Chemistry – unpreserved plastic bottle (also applicable)

Preservative: Refrigerate or store on ice. Do not allow to freeze.


Handling: **Recommend next day receipt at lab.** Sample testing must begin within 28 days after sample collection.

Method Technology: Specific Conductance

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Corrosivity (Langelier Index)
Method: SM 2330B



Note: Corrosivity requires additional testing for Calcium, TDS, Alkalinity as CaCO_3 , and pH.

Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container. If preservative is present, be sure not to over fill to prevent loss of preservative

Required Containers/Volume: 250 mL Total Metals – plastic bottle
1 L Total Chemistry – plastic bottle

Preservative: For Total Metals bottle, add HNO_3 to pH <2. Refrigerate or store on ice. Do not allow to freeze.

Handling: **Recommend next day receipt at lab.** Total Metals sample testing must begin within 180 days after collection. Total Chemistry sample testing must begin within 48 hours after collection.

Method Technology: Calculation using Calcium, TDS, Alkalinity, and pH results

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Cyanide [CN⁻]
Method: EPA 335.4

Application: Drinking Water, Surface Water,
Groundwater

Analytes: Cyanide



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container, be sure not to over fill to prevent loss of preservative.

Required Containers/Volume: 500 mL Cyanide – plastic bottle

Preservative: 2 g NaOH to pH >12, ascorbic acid in the presence of residual chlorine. Refrigerate or store on ice. Do not allow to freeze.

Handling: **Recommend next day receipt at lab.** Sample testing must begin within 14 days after collection.

Method Technology: Flow Injection Colorimetry

Preparation Method: Distillation required before analysis

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Fluoride [F⁻]
Method: EPA 300.0

Application: Drinking Water, Surface Water

Analytes: Fluoride



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: 120 mL Fluoride – unpreserved plastic bottle

Preservative: Refrigerate or store on ice. Do not allow to freeze.

Handling: **Recommend next day receipt at lab.** Sample testing must begin within 28 days of sample collection.

Method Technology: Ion Chromatography

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Nitrate + Nitrite
[NO₃⁻]+[NO₂⁻],
Nitrite [NO₂⁻]

Method: EPA 353.2

Groupings: Total Nutrients
Filtered Nutrients

Application: Drinking Water, Surface Water

Analytes: Nitrate/Nitrite (NO₃+NO₂)
Nitrite (NO₂)

Note: Prior to collection of Nitrate and Nitrite, analysis must be scheduled at (801) 965-2400.



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container, be sure not to over fill to prevent loss of preservative.

Required Containers/Volume: 120 mL Nitrate – plastic bottle
500 mL Total Nutrients – plastic bottle (also applicable)
120 mL Nitrite – unpreserved plastic bottle

Preservative: **Nitrate** – H₂SO₄ to pH <2, refrigerate or store on ice. Do not allow to freeze.
Nitrite – No preservative, refrigerate or store on ice. Do not allow to freeze.

Handling: Nitrate – **Recommend next day receipt at lab.** Sample testing must begin within 28 days after collection.
Nitrite – **Recommend next day receipt at lab.** Sample testing must begin within 48 hours after collection.

Method Technology: Flow Injection Colorimetry

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Nitrogen, Total
Method: ASTM D 8083

Groupings: Filtered Nutrients

Application: Drinking Water, Surface Water

Analytes: Nitrogen



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container, be sure not to over fill to prevent loss of preservative

Required Containers/Volume: 250 mL Filtered Nutrient-plastic bottle

Preservative: H_2SO_4 to pH <2, refrigerate or store on ice. Do not allow to freeze.

Handling: Recommend next day receipt at lab. Sample testing must begin within 28 days of sample collection.

Method Technology: High temperature catalytic combustion & chemiluminescence detection

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Odor
Method: EPA 140.1

Groupings: New Drinking Water Source

Application: Drinking Water, Surface Water

Analytes: Odor (Threshold Odor Number, TON)

Note: Prior to collection of Odor, analysis must be scheduled at (801) 965-2400.



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: 1 L Odor – amber glass bottle

Preservative: Refrigerate or store on ice. Do not allow to freeze.

Handling: Recommend same day receipt at lab. Sample testing must begin within 24 hours after sample collection.

Method Technology: Odor Threshold (Consistent Series)

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Perchlorate [ClO₄]
Method: Subcontracted

NOTE: This test is NOT performed at UPHL and will be sent to a subcontracted laboratory for testing.

Application: Drinking Water

Analytes: Perchlorate
Conductivity



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: 120 mL Perchlorate – unpreserved plastic bottle
1L Total Chemistry – unpreserved plastic bottle (also applicable)

Preservative: Refrigerate or store on ice. Do not allow to freeze.

Handling: **Recommend next day receipt at lab.** Sample testing must begin within 28 days after sample collection.

Method Technology: Ion Chromatography

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: pH
Method: EPA 150.1

Groupings: Total Chemistry

Application: Drinking Water, Surface Water

Analytes: pH (Lab Measured)



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: 1 L Total Chemistry – unpreserved plastic bottle

Preservative: Refrigerate or store on ice. Do not allow to freeze.

Handling: **Recommend same day receipt at lab.** Sample testing must begin within 24 hours after sample collection.

Method Technology: Electrometric Measurement

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Phosphorus (All Forms)
Method: EPA 365.1

Groupings: Total Nutrients
Dissolved Nutrients

Application: Drinking Water, Surface Water

Analytes: Phosphorus, Total (All Forms)



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container, be sure not to over fill to prevent loss of preservative.

Required Containers/Volume: 500 mL Nutrient Analysis – plastic bottle

Preservative: H_2SO_4 to pH <2, refrigerate or store on ice. Do not allow to freeze.

Handling: Recommend same day receipt at lab. Sample testing must begin within 28 days after sample collection.

Method Technology: Flow Injection Colorimetry

Preparation Method: Digestion technique

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: ORTHO Phosphate

[PO₄³⁻]

Method: EPA 300.1

Groupings: Total Nutrients
Dissolved Nutrients

Application: Drinking Water, Surface Water

Analytes: Phosphate, ORTHO

Note: Prior to collection of Phosphate (Ortho), analysis must be scheduled with the laboratory at (801) 965-2400.



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container, be sure not to over fill to prevent loss of preservative.

Required Containers/Volume: 500 mL Nutrient Analysis – plastic bottle

Preservative: H₂SO₄ to pH <2, refrigerate or store on ice. Do not allow to freeze.

Handling: Recommend same day receipt at lab. ORTHO-Phosphate has a very short holding time, please notify the laboratory to coordinate sample submission.

Method Technology: Ion Chromatography

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Silica [SiO_2]
Method: SM 4500-SiO₂ F

Application: Drinking Water, Surface Water

Analytes: Silica (SiO_2)



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: 1 L Total Chemistry – unpreserved plastic bottle

Preservative: Refrigerate or store on ice. Do not allow to freeze.

Handling: **Recommend same day receipt at lab.** Sample testing must begin within 28 days after sample collection.

Method Technology: Flow Injection Colorimetry

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Sulfate [SO_4^{2-}]
Methods: EPA 300.0, EPA 375.2

Application: Drinking Water (300.0)
Surface Water (375.2)



Analytes: Sulfate

Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: 120 mL Sulfate – unpreserved plastic bottle

Preservative: Refrigerate or store on ice. Do not allow to freeze.

Handling: **Recommend next day receipt at lab.** Sample testing must begin within 28 days after sample collection.

Method Technology: Ion Chromatography and Flow Injection Colorimetry

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Sulfide [S^{2-}]
Method: SM 4500- S^{2-} D/
HACH Method 8131

Application: Drinking Water, Surface Water

Analytes: Sulfide



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container, be sure not to over fill to prevent loss of preservative.

Required Containers/Volume: 120 mL Sulfide – plastic bottle

Preservative: 3 drops Zinc Acetate and NaOH to pH >9. Refrigerate or store on ice. Do not allow to freeze.

Handling: Recommend next day receipt at lab. Sample testing must begin within 7 days after sample collection.

Method Technology: Colorimetry

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Solids
Methods: SM 2540C-TDS, SM 2540D-TSS, SM 2540E-VSS

Application: Drinking Water, Surface Water



Analytes: Solids, Total Dissolved (TDS) – filterable
Solids, Total Suspended (TSS) – non-filterable
Solids, Volatile Suspended (VSS)

Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: 1 L Total Chemistry – unpreserved plastic bottle

Preservative: Refrigerate or store on ice. Do not allow to freeze.

Handling: **Recommend same day receipt at lab.** Sample testing must begin within 7 days after collection, except SS which is 48 hours.

Method Technology: Gravimetric detection

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Turbidity
Method: EPA 180.1

Groupings: Total Chemistry
Water Treatment Technique

Application: Drinking Water, Surface Water

Analytes: Turbidity (NTU)



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: 1 L Total Chemistry – unpreserved plastic bottle

Preservative: Refrigerate or store on ice. Do not allow to freeze.

Handling: **Recommend same day receipt at lab.** Sample testing must begin within 48 hours after sample collection.

Method Technology: Nephelometric Absorbance

[Back to Inorganic Analytical Services](#)

Metals & Inorganic Chemistry

Name: Annual Inorganics & Metals (18 Parameters)
Methods: Type 9 (Primary Inorganics & Metals Chemistry)



Analytes: Cyanide	Fluoride
Turbidity (NTU)	Solids (TDS)
T-Arsenic	T-Barium
T-Beryllium	T-Cadmium
T-Chromium	T-Copper
T-Lead	T-Mercury
T-Nickel	T-Selenium
T-Antimony	T-Thallium
T-Sodium	Sulfate

Instructions for Collection: Allow sample tap to flow for a few minutes until water temperature stabilizes. Slowly fill containers to top, be sure not to over fill to prevent loss of preservative.

Required Containers/Volume: 1 L Total Chemistry – unpreserved plastic bottle
250 mL Total Metals – plastic bottle
500mL Cyanide – plastic bottle

Preservative: Refrigerate or store on ice. Do not allow to freeze.

Handling: Recommend same day receipt at lab. Must be received at lab within 24 hours of collection time.

Method Technology: Nephelometric, Gravimetric, Flow Injection Colorimetry, Ion Chromatography, ICP, and ICPMS

[Back to Inorganic Analytical Services](#)

Metals & Inorganic Chemistry

Name: New Drinking Water Source (46 Parameters)

Method: Type 7 (Total Inorganics & Metals Chemistry)

Analytes

Odor (TON)	Solids (TDS)	T-Arsenic
Surfactant	Solids (TSS)	T-Barium
Alkalinity	Turbidity (NTU)	T-Beryllium
Bicarbonate	Cyanide	T-Boron
Carbon Dioxide	Sulfate	T-Cadmium
Carbonate	Ammonia	T-Chromium
Carbonate Solids	NO ₂ +NO ₃	T-Copper
Chloride	T-Phosphate	T-Iron
Corrosivity	Color	T-Lead
Fluoride	D-Calcium	T-Manganese
Hardness	D-Magnesium	T-Mercury
Hydroxide	D-Potassium	T-Selenium
L-pH or F-pH	D-Sodium	T-Silver
L-Specific Conductivity	T-Aluminum	T-Thallium
Silica	T-Antimony	T-Zinc
		Total Organic Carbon (TOC)



Instructions for Collection:

Note: Prior to collection of New Drinking Water Source samples, analysis must be scheduled with the laboratory at (801) 965-2400.

Allow sample tap to run for a few minutes until water temperature stabilizes. Slowly fill to top of bottles, be sure not to over fill to prevent loss of preservatives.

Required Containers/Volume:

- 1 L Odor – amber glass bottle
- 1 L Surfactant – amber glass bottle
- 1 L Total Chemistry – unpreserved plastic bottle
- 500 mL Cyanide – plastic bottle
- 500 mL Nutrient Analysis – plastic bottle
- 250 mL Color – plastic bottle
- *250 mL Filtered Metals – plastic bottle
- *250 mL Total Metals – plastic bottle
- 200 mL TOC – amber glass bottle

Preservative:

Refrigerate or store on ice. Do not allow to freeze. **Metals Preserve with HNO₃ to pH <2.**

Handling:

Recommend same day receipt at lab. Must be received at lab within 24 hours of collection time.

Method Technology:

Electrometric, Nephelometric, Gravimetric, Titration, Flow Injection Colorimetry, Ion Chromatography, ICP, and ICPMS

[Back to Inorganic Analytical Services](#)

Metals & Inorganic Chemistry

Name: Total Chemistry Analytical Groupings



Type 2 or 3



Type 9

L = Lab Measured
 D = Dissolved
 T = Total

Type 2

pH
 Solids (TSS)
 Bicarbonate
 Carbon Dioxide
 Carbonate
 Chloride
 Hydroxide
 Sulfate
 Alkalinity
 Turbidity (NTU)
 Conductivity
 Solids (TDS)

Type 3

pH
 Solids (TSS)
 D-Calcium
 D-Magnesium
 D-Potassium
 Bicarbonate
 Carbon Dioxide
 Carbonate
 Chloride
 Hydroxide
 Sulfate
 Alkalinity
 Hardness
 Turbidity (NTU)
 Conductivity
 Solids (TDS)
 Carbonate Solids

Type 9

Cyanide
 Turbidity (NTU)
 T-Arsenic
 T-Beryllium
 T-Chromium
 T-Lead
 T-Nickel
 T-Antimony
 T-Sodium
 Fluoride
 Solids (TDS)
 T-Barium
 T-Cadmium
 T-Copper
 T-Mercury
 T-Selenium
 T-Thallium
 Sulfate

Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: 1 L Total Chemistry – unpreserved plastic bottle
 1 L Total Chemistry – unpreserved plastic bottle
 250 mL Total Metals – plastic bottle
 500mL Cyanide – plastic bottle

Preservative: Refrigerate or store on ice. Do not allow to freeze.

Handling: **Recommend next day receipt at lab.** Sample testing must begin within 48 hours after sample collection.

Method Technology: Electrometric, Nephelometric, Gravimetric, Titration, Flow Injection Colorimetry, Ion Chromatography, ICP, and ICPMS

[Back to Inorganic Analytical Services](#)

Inorganic Chemistry

Name: Nutrients, Total & Dissolved Analytical Groupings

Type 2

Ammonia

Phosphate (T-PO4)

Type 6

Phosphate (T-PO4)

Nitrate+Nitrite (NO₃+NO₂)

Type 3

Ammonia

Phosphate (T-PO4)

Nitrate+Nitrite (NO₃+NO₂)

Type 9

Phosphate (D-PO4)

D-Nitrate+Nitrite (NO₃)

D-Total Nitrogen



Type 2, 3 or 6



Type 9

Note: To test for Dissolved analytes, field filtration must be done at time of sample collection and Filtered-Nutrients bottle used.

D = Dissolved

T = Total

Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container, be sure not to over fill to prevent loss of preservative.

Required Containers/Volume: 500 mL Nutrient Analysis – plastic bottle
250 mL Filtered Nutrients – plastic bottle

Preservative: H₂SO₄ to pH <2. Refrigerate or store on ice. Do not allow to freeze.

Handling: **Recommend next day receipt at lab.** Sample testing must begin within 28 days after sample collection.

Method Technology: Flow Injection Colorimetry

Preparation Method: Digestion Techniques (except for Nitrate + Nitrite)

[Back to Inorganic Analytical Services](#)

Metals

Analytical Methods for Metals

EPA 200.8	Drinking Water and Groundwater
EPA 200.7	Drinking Water, Groundwater, and Wastewater
EPA 6010	Wastes (SW 846), Water, Soil
EPA 6020	Wastes (SW 846), Water, Soil
EPA 245.1	Mercury in Water (Prep Method 245.1)
EPA 7471B	Mercury in Soil (Prep Method 7471B)

Digestion Methods for Metals (Non-Drinking Water)

EPA 3010A	TCLP & Total Metal Sample Digestion
EPA 200.8	Waste water, Groundwater and Drinking Water
EPA 3050B	Solid and Hazardous Wastes

Note: Drinking water samples generally do not require digestion unless water exceeds turbidity of 1 NTU.

[Back to Metals Analytical Services](#)

Metals

Name: Arsenic [As]
Method: EPA 200.8

Groupings: Metals

Application: Arsenic Rule, Drinking Water, Surface Water

Analytes: Arsenic (As, Total & Dissolved)



Instructions for Collection: Allow sample tap to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container, be sure not to over fill to prevent loss of preservative.

Required Containers/Volume: 250 mL Total Metals – plastic bottle

Preservative: HNO₃ to pH <2, refrigerate or store on ice. Do not allow to freeze.

Handling: Recommend next day receipt at lab. Sample must be analyzed within 6 months after collection.

Method Technology: Digestion technique followed by ICPMS detection

[Back to Metals Analytical Services](#)

Metals

Name: Mercury [Hg]
Method: EPA 245.1, EPA 7471B

Groupings: Metals

Application: Drinking Water, Surface Water, Soil, Solids, Sludge

Analytes: Mercury (Hg, Total & Dissolved)



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container, be sure not to over fill to prevent loss of preservative.

Required Containers/Volume: **Water** – 250 mL Total Metals – plastic bottle
Soils, solids, sludge – 4 oz. glass container with Teflon-lined lid

Preservative: HNO₃ to pH <2. Refrigerate or store on ice (4-6°C). Do not allow to freeze.

Handling: **Recommend next day receipt at lab.** Sample testing must begin within 28 days after collection. If sample is not preserved, sample testing must begin within 24 hrs.

Method Technology: Cold Vapor AA detection

Preparation Method: Digestion Technique

[Back to Metals Analytical Services](#)

Metals

Name: Chromium-VI [Cr^{6+}]
Method: Subcontracted

NOTE: This test is NOT performed at UPHL and will be sent to a subcontracted laboratory for testing.

Application: Drinking Water, Surface Water

Analytes: Chromium-VI (Hexavalent)



Note: Prior to collection of Chromium-VI samples, analysis must be scheduled at (801) 965-2400.

Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: 120 mL Chromium-VI – plastic bottle

Preservative: 1 mL of $(\text{NH}_4)_2\text{SO}_4/\text{NH}_4\text{OH}$ to 100 mL. Refrigerate or store on ice. Do not allow to freeze.

Handling: Recommend same day receipt at lab. Sample testing must begin within 14 days after sample collection.

Method Technology: Ion Chromatography

[Back to Metals Analytical Services](#)

Metals

Name: Lead & Copper
Method: EPA 200.8

Groupings: Metals

Application: Corrosion Control Assessment,
Drinking Water

Analytes: Lead (Pb, Total)
Copper (Cu, Total)



Instructions for Collection: Allow sample tap to flow for a few minutes until water temperature stabilizes. Slowly fill to line.

Required Containers/Volume: 1 L Lead and Copper – plastic bottle

Preservative: Refrigerate or store on ice. Do not allow to freeze.
Sample must be preserved with HNO₃ to pH <2 within
14 days after sample collection.

Handling: **Recommend next day receipt at lab.** If preserved
within 14 days after sample collection, testing must
begin within 6 months.

Method Technology: ICPMS Detection

Preparation Method: Digestion Technique

[Back to Metals Analytical Services](#)

Metals

Name: Hardness (Hardness as CaCO_3)

Method: SM 2340B

Application: Drinking Water, Surface Water

Analytes: Calcium (Ca)
Magnesium (Mg)



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container, be sure not to over fill to prevent loss of preservative.

Required Containers/Volume: 250 mL Total Metals – plastic bottle

Preservative: Nitric acid (HNO_3) to pH <2. Refrigerate or store on ice. Do not allow to freeze.

Handling: Store filtered samples in plastic bag

Method Technology: ICP-MS Detection of Ca & Mg. Calculation: Based on calcium and magnesium levels

Preparation Method: Digestion is necessary if turbidity is >1 NTU.

[Back to Metals Analytical Services](#)


Metals

Name: Toxic Characteristic Leaching Procedure (TCLP) RCRA

Method: TCLP-Metals, 1311-TCLP, 1332-TCLP

Groupings: Metals Testing (RCRA)

Note: Prior to collection of 1311-TCLP (water and soil) or 1332-TCLP (oil), analysis must be scheduled at (801) 965-2400.



*Other metals may be analyzed, but must be specified on test request form. For complete list of metals see indexes or contact us 801-965-2400.

<u>HW</u>	<u>RCRA 8</u>	<u>RCRA 8+4</u>	<u>RCRA 8+4+6</u>
D004	T-Arsenic	T-Arsenic	T-Arsenic
D005	T-Barium	T-Barium	T-Barium
D006	T-Cadmium	T-Cadmium	T-Cadmium
D007	T-Chromium	T-Chromium	T-Chromium
D008	T-Lead	T-Lead	T-Lead
D009	T-Mercury	T-Mercury	T-Mercury
D010	T-Selenium	T-Selenium	T-Selenium
D011	T-Silver	T-Silver	T-Silver
		T-Copper	T-Copper
		T-Iron	T-Iron
		T-Manganese	T-Manganese
		T-Zinc	T-Zinc
			T-Aluminum
			T-Beryllium
			T-Cobalt
			T-Molybdenum
			T-Nickel
			T-Vanadium

Instructions for Collection: Slowly fill to top of container.

Required Containers/Volume: **Water** - 4 L amber glass bottle

Soil - 4 oz. amber with Teflon-lined lid glass container

Preservative: Refrigerate or store on ice 4-6°C. Do not allow to freeze.

Handling: **Recommend next day receipt at lab.** Samples must be TCLP extracted within 7 days after collection and must be analyzed within 180 days, except Mercury which must be analyzed within 28 days after collection.


Method Technology: Method 6010 ICP, Method 6020 ICPMS, and Cold Vapor AA Detections

Preparation Method: Method 1311 Leaching Procedure and Extraction Technique

[Back to Metals Analytical Services](#)

Metals

Name: Total Metals & Dissolved Metals Analytical Groupings



Note: To test for dissolved analytes, field filtration must be done at time of sample collection and a filtered-metals bottle used. The constituents of the types are subject to change based upon customer request.

Type 3

D-Aluminum
D-Arsenic
D-Barium
D-Boron
D-Cadmium
D-Calcium
D-Chromium
D-Copper
D-Iron
D-Lead
D-Magnesium
D-Manganese
D-Mercury
D-Nickel
D-Potassium
D-Selenium
D-Silver
D-Sodium
D-Zinc
Hardness

Type 4

D-Calcium
D-Magnesium
D-Potassium
D-Sodium
Hardness

Type 7

T-Aluminum
T-Arsenic
T-Barium
T-Boron
T-Cadmium
T-Calcium
T-Chromium
T-Copper
T-Iron
T-Lead
T-Magnesium
T-Manganese
T-Mercury
T-Nickel
T-Potassium
T-Selenium
T-Silver
T-Sodium
T-Zinc

Type 9

T-Barium
T-Cadmium
T-Chromium
T-Mercury
T-Selenium

RCRA 8

T-Arsenic
T-Barium
T-Cadmium
T-Chromium
T-Lead
T-Mercury
T-Selenium
T-Silver

RCRA 8+4

T-Arsenic
T-Barium
T-Cadmium
T-Chromium
T-Lead
T-Mercury
T-Selenium
T-Silver
T-Copper
T-Iron
T-Manganese
T-Zinc

RCRA 8+4+6

T-Arsenic
T-Barium
T-Cadmium
T-Chromium
T-Lead
T-Mercury
T-Selenium
T-Silver
T-Copper
T-Iron
T-Manganese
T-Zinc
T-Aluminum
T-Beryllium
T-Cobalt
T-Molybdenum
T-Nickel
T-Vanadium

Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container, be sure not to over fill to prevent loss of preservative.

Required Containers/Volume: 250 mL Total Metals – plastic bottle

250 mL Filtered Metals – plastic bottle
Soils, solids, sludge– 4 oz. glass container with Teflon-lined lid

Preservative: Liquids add HNO₃ to pH <2. Refrigerate or store on ice. Do not allow to freeze.

Handling: **Recommend next day receipt at lab. Sample testing must begin within 180 days after collection, except Mercury which is 28 days (24 hrs for Hg if not preserved).**

Method Technology: ICP (EPA 200.7, 6010), ICPMS (EPA 200.8, 6020), and Cold Vapor AA Detections (EPA 245.1)

Preparation Method: Digestion Techniques (except for Drinking Water)

[Back to Metals Analytical Services](#)

Organic Chemistry

Name: Chlorophyll-A &
Pheophytin-A

Method: SM 10200H (modified)

Application: Surface Water

Analytes: Chlorophyll-A
Pheophytin-A



Instructions for Collection: Filter up to 1000 mL of water through a glass fiber filter and place filter in an opaque container. **Record the volume of sample filtered on the test request form.**

Required Containers/Volume: Glass fiber filters; Store filter in an opaque container. Use one filter per sample.

Preservative: Freeze sample filter.

Handling: Keep frozen. Protect from light. **Recommend next day receipt at lab.** Must analyze sample within 28 days after collection.

Method Technology: UV-VIS Spectrophotometry or HPLC

Preparation Method: Homogenization; addition of acid for UV-VIS Pheophytin

[Back to Organic Analytical Services](#)

Organic Chemistry

Name: Cyanotoxins
Method: EPA 546 & EPA 546
Modified, ELISA

Application: Drinking Water, Surface Water



*Microcystins &
Cylindrosperm.*



Anatoxin-a

Analytes: Anatoxin-A
Cylindrospermopsin
Total Microcystins & Nodularins

Instructions for Collection: Carefully fill to shoulder of bottle. Do not immerse or overfill bottle or preservative will be lost.

Required Containers/Volume: 250-mL amber glass bottle. One bottle for Cylindrospermopsin & Microcystins. One bottle for Anatoxin-a.

Preservative: **Surface Water:** Micro/Cylin bottle – no preservative. Anatoxin bottle – 250 mg sodium bisulfate.

Drinking Water: Micro/Cylin bottle – 25 mg ascorbic acid & 25 mg sodium thiosulfate. Anatoxin bottle – 250 mg sodium bisulfate & 25 mg ascorbic acid.

Samples may be stored refrigerated up to 5 days after collection. After 5 days, the samples should be stored frozen.

Handling: Keep samples refrigerated or store on ice. **Recommend same day receipt at lab.** Sample prep and analysis should begin as soon as possible upon receipt.

Method Technology: Enzyme-linked immunosorbent assay (ELISA)

Preparation Method: Freeze/Thaw Cycles

[Back to Organic Analytical Services](#)

Organic Chemistry

Name: Haloacetic Acids (HAAs)
Method: SM 6251B
/Subcontracted

NOTE: This test is NOT performed at UPHL and will be sent to a subcontracted laboratory for testing



Grouping: Disinfection By-Products

Application: Water systems using chlorine or bromine for disinfection.

Analytes: Dibromoacetic Acid (DBAA)
Dichloroacetic Acid (DCAA)
Monobromoacetic Acid (MBAA)
Monochloroacetic Acid (MCAA)
Trichloroacetic Acid (TCAA)

Instructions for Collection: Allow sample tap to run for a few minutes until water temperature stabilizes. Slowly fill vials to top of container, be sure not to over fill to prevent loss of preservative. Check for air bubbles by inverting. Fill the remaining space if bubbles are observed. **There must be no headspace.**

Required Containers/Volume: 3/40 mL vials

Preservative: 65 mg ammonium chloride (NH_4Cl)

Handling: Refrigerate or store on ice. Do not allow sample to freeze. **Recommend next day receipt at lab.** Sample must be extracted within 14 days of sample collection.

Method Technology: Analysis by GC-ECD

Preparation Method: Liquid-liquid Extraction

[Back to Organic Analytical Services](#)

Organic Chemistry

Name: Organic Carbon, Total (TOC) & Dissolved (DOC)

Method: SM 5310B

Grouping: Total Nutrients, Filtered Nutrients

Application: Drinking Water, Surface Water, Groundwater

Analytes: Organic Carbon, Total (TOC)
Organic Carbon, Dissolved (DOC)



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to the top of the container. Be sure not to over fill to prevent loss of preservative.

Required Containers/Volume: 200 mL TOC – amber glass bottle

Preservative: H₂SO₄ to pH <2. Refrigerate or store on ice. Do not allow it to freeze.

Handling: Recommend next day receipt at lab. Sample must be analyzed within 28 days after collection.

Method Technology: Combustion / Infrared Detection

[Back to Organic Analytical Services](#)

Organic Chemistry

Name: Organic Constituents,
UV-Absorbing
Method: SM 5910B
/Subcontracted

NOTE: This test is NOT performed at UPHL and will be sent to a subcontracted laboratory for testing

Grouping: Total Nutrients
Water Treatment Technique

Application: Drinking Water, Surface Water,
Groundwater

Analytes: UV-Absorbing Organics



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: 200 mL UV254 – amber glass bottle

Preservative: Refrigerate or store on ice. Do not allow to freeze.

Handling: Recommend next day receipt at lab. Sample analysis must begin within 48 hours after sample collection.

Method Technology: UV Absorbance

[Back to Organic Analytical Services](#)

Organic Chemistry

Name: Periphyton
Method: SM 10300C (modified)

Application: Surface Water



Analytes: Periphyton

Instructions for Collection: Follow sample collector's protocol or SM10200H (modified)

Required Containers/Volume: Glass Fiber filter paper– store in opaque container or follow sample collector's instructions

Preservative: Keep frozen.

Handling: **Recommend next day receipt at lab.** Sample testing must begin as soon as possible after sample collection.

Method Technology: Gravimetry

Preparation Method: Dry-Ashing Technique

[Back to Organic Analytical Services](#)

Organic Chemistry

Name: Trihalomethanes (THMs),
Max Potential THMs
Method: EPA 524.2-THM
/Subcontracted

NOTE: This test is NOT performed at UPHL and will be sent to a subcontracted laboratory for testing.



Grouping: Disinfection By-Products

Application: Water systems using chlorine or bromine for disinfection.

Analytes: Bromodichloromethane
Bromoform
Chlorodibromomethane
Chloroform
Maximum Potential THMs, EPA 510.1

Instructions for Collection: Allow sample tap to run for a few minutes until water temperature stabilizes. Slowly fill vials to top of container, be sure not to over fill to prevent loss of preservative. Check for air bubbles by inverting. Fill the remaining space if bubbles are observed. **There must be no headspace.**

Required Containers/Volume: 3/40 mL vials

Preservative: 4 mg $\text{Na}_2\text{S}_2\text{O}_3$

Handling: Refrigerate or store on ice. Do not allow sample to freeze. **Recommend next day receipt at lab.** Sample must be analyzed within 14 days of sample collection.

Method Technology: Purge and trap technique; Analysis by GC-MS

[Back to Organic Analytical Services](#)

Organic Chemistry

Name: BTEXN
Method: EPA 8260B
/Subcontracted

NOTE: This test is NOT performed at UPHL and will be sent to a subcontracted laboratory for testing.

Application: Wastewater, Soil

Analytes: Benzene
Toluene
Ethylbenzene
Xylenes
Naphthalene

Note: If using Encore sampler, please bring to lab as soon as possible (within 48 hours).



Instructions for Collection: If applicable, allow sample tap to run for a few minutes until water temperature stabilizes. Slowly fill vials to top of container. Check for air bubbles by inverting, and fill the remaining portion if bubbles found. There must be no headspace.

Required Containers/Volume: Water – 3/40 mL vials
Soil – 4 oz. glass container with Teflon-lined lid

Preservative: Refrigerate or store on ice. Do not allow to freeze.

Handling: Recommend next day receipt at lab. Sample testing must begin within 14 days after sample collection.

Method Technology: Purge and trap technique followed by GCMS detection

Preparation Method: Water-EPA 5030; Soil-EPA 3585

[Back to Organic Analytical Services](#)

Organic Chemistry

Name: Carbamates (Insecticides & Aldicarbs)

Method: EPA 531.1
/Subcontracted

NOTE: This test is NOT performed at UPHL and will be sent to a subcontracted laboratory for testing.

Application: Drinking Water

Analytes: 3-Hydroxycarbofuran
Aldicarb (Temik)
Aldicarb Sulfone
Aldicarb Sulfoxide
Carbaryl (Sevin)
Carbofuran (Furadan)
Methomyl
Oxamyl (Vydate)



Instructions for Collection: Allow sample tap to run for a few minutes until water temperature stabilizes. Slowly fill vial to top of container, be sure not to over fill to prevent loss of preservative. Check for air bubbles by inverting, and fill the remaining portion if bubbles found. There must be no headspace.

Required Containers/Volume: 40 mL amber glass vial

Preservative: 1.2 mL monochloroacetic acid. Refrigerate or store on ice. Do not allow to freeze.

Handling: Recommend next day receipt at lab. Sample testing must begin within 28 days after sample collection.

Method Technology: Aqueous Injection HPLC with Post-column Derivation and Fluorescence Detection

[Back to Organic Analytical Services](#)

Organic Chemistry

Name: Geosmin & MIB (Odor)
Method: EPA 525.2
/Subcontracted

NOTE: This test is NOT performed at UPHL and will be sent to a subcontracted laboratory for testing.

Application: Drinking Water

Analytes: Geosmin
2-Methylisoborneol (MIB)



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: 1 L Method EPA 525.2 Odor – amber glass bottle

Handling: Same day receipt at lab. No holding time

Method Technology: GCMS Detection

Preparation Method: Liquid-solid extraction

[Back to Organic Analytical Services](#)

Organic Chemistry

Name: Herbicides (Chlorinated Organic Acids)

Method: EPA 515.1, EPA 8151 /Subcontracted

NOTE: This test is NOT performed at UPHL and will be sent to a subcontracted laboratory for testing.

Application: Drinking Water, Surface Water, Wastewater, Groundwater, Soil



Analytes: 2,4-D
2,4,5-TP (Silvex)
Dalapon
Dicamba
Dinoseb
Pentachlorophenol
Picloram

Instructions for Collection: If applicable, allow sample tap to run for a few minutes until water temperature stabilizes. Slowly fill bottles to top of container, be sure not to over fill to prevent loss of preservative. Collect per project sampling plan.

Required Containers/Volume: **Water** – 2/1L amber glass bottles
Soils, Solids, and Sludge– 4 oz. amber glass container with Teflon-lined lid

Preservative: 30 mg $\text{Na}_2\text{S}_2\text{O}_3$ if chlorinated. Refrigerate or store on ice. Do not allow to freeze.

Handling: **Recommend next day receipt at lab. Sample extraction must begin within 14 days after collection. Analysis must begin within 28 days after extraction.**

Method Technology: GC-ECD Detection

Preparation Method: EPA 515.1 and 8151 (Water)- Liquid-liquid extraction; EPA 8151-(Soils) Solid-liquid extraction, then liquid-liquid extraction.

[Back to Organic Analytical Services](#)

Organic Chemistry

Name: Oil and Grease
Method: EPA 1664 /Subcontracted

NOTE: This test is NOT performed at UPHL and will be sent to a subcontracted laboratory for testing.



Analytes: Total O/G (Oil and Grease)

Instructions for Collection: If applicable, allow source to flow for a few minutes until water temperature stabilizes. Slowly fill containers to top, be sure not to over fill to prevent loss of preservative.

Required Containers/Volume: Water – 1 L glass container with Teflon-lined lid
Soil – 4 oz. glass container with Teflon-lined lid

Preservative: H_2SO_4 to pH <2. Refrigerate or store on ice. Do not allow to freeze.

Handling: Recommend next day receipt at lab. Sample testing must begin within 28 days after collection.

Method Technology: Gravimetric Detection

Preparation Method: Liquid-solid Extraction

[Back to Organic Analytical Services](#)

Organic Chemistry

Name: Pesticides & Semi-Volatile Organic Compounds (SVOCs)
Method: EPA 525.2, EPA 625, EPA 8270 /Subcontracted

NOTE: This test is NOT performed at UPHL and will be sent to a subcontracted laboratory for testing.

Application: Drinking Water, Surface Water, Wastewater, Groundwater, Solids

Analytes: Refer to Program Specific Indexes



CWA: [Page 67](#)

RCRA: [Page 72](#)

SDWA: [Page 77](#)

Instructions for Collection: For drinking water samples, pour the small vial of acid into each sample bottle. Allow source to flow for a few minutes until water temperature stabilizes, **do not use Tygon tubing**. Slowly fill bottles to top of container, be sure not to over fill to prevent loss of preservative.

Required Containers/Volume: **Drinking Water** – 2/1L amber glass bottles + acid vials
Surface Water – 2/1L amber glass bottles
Groundwater – 2/1L amber glass bottles
Soil, solids, sludges – 4 oz. amber glass container with Teflon-lined lid

Preservative: **Drinking Water** – HCl to pH <2, 50 mg sodium sulfite
Surface Water – No preservative
Groundwater and Solids – No preservative

All sample types, refrigerate or store on ice 4-6°C. Do not allow to freeze.
Handling: **Recommend next day receipt at lab.** Sample extraction/analysis times: EPA 525.2 – 14 days to extract, then 30 days to analyze. EPA 625 and EPA 8270 – 7 days to extract, then 40 days to analyze. Solids – 14 days to extract, then 40 days to analyze.

Method Technology: GCMS Detection

Preparation Method: Liquid-solid Extraction; Groundwater (EPA 8270)-EPA 3510; Soil (EPA 8270)-EPA 3550 or EPA 3545A

[Back to Organic Analytical Services](#)

Organic Chemistry

Name: Polychlorinated
Biphenyls (PCBs) &
Organochlorine
Pesticides

Method: EPA 608, EPA 8081,
EPA 8082 /Subcontracted

NOTE: This test is NOT performed at UPHL and will be sent to a subcontracted laboratory for testing.

Application: Surface Water, Groundwater,
Wastewater, Solids

Analytes: Refer to Program Specific Indexes

CWA: [Page 67](#)

RCRA: [Page 72](#)



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: **Surface Water** – 1 L amber glass bottle
Groundwater – 1 L amber glass bottle
Soils, solids, sludge – 4 oz. amber glass container with Teflon-lined lid
Oil – 4 oz. amber glass container with Teflon-lined lid

Preservative: Refrigerate or store on ice. Do not allow to freeze.

Handling: **Recommend next day receipt at lab.** Sample extraction must begin within 7 days of collection (except Soil – 14 days). Analysis must begin within 40 day of extraction.

Method Technology: GCMS detection

Preparation Method: Extraction technique (Liquid-liquid, sonication, waste dilution);
Water- EPA 3510; Oils-EPA 3580; Soil-EPA 3550.

[Back to Organic Analytical Services](#)

Organic Chemistry

Name: Surfactants
Method: SM 5540C
/Subcontracted

NOTE: This test is NOT performed at UPHL and will be sent to a subcontracted laboratory for testing.

Grouping: New Drinking Water Source

Application: Drinking Water, Surface Water

Analytes: Total Surfactants



Instructions for Collection: Allow sample tap to run for a few minutes until water temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: 1 L amber glass bottle

Preservative: Refrigerate or store on ice. Do not allow to freeze.

Handling: Recommend same day receipt at lab. Sample testing must begin within 48 hours after sample collection.

Method Technology: MBAS Detection

Preparation Method: Extraction Technique

[Back to Organic Analytical Services](#)

Organic Chemistry

Name: Total Petroleum
Hydrocarbons (TPH)
Method: EPA 8015B
/Subcontracted

NOTE: This test is NOT performed at UPHL and will be sent to a subcontracted laboratory for testing.

Application: Surface Water, Wastewater,
Groundwater, Soil

Analytes: GRO-Gasoline Range Organics (C6-C10)
DRO-Diesel Range Organics (C10-C28)
ORO-Oil Range Organics (C28-C35)

Total TPH = GRO+DRO+ORO



Instructions for Collection: If applicable, allow sample tap to run for a few minutes until water temperature stabilizes. Slowly fill vials to top of container. Check for air bubbles by inverting, and fill the remaining portion if bubbles found. There must be no headspace.

Required Containers/Volume: **Water** – 3/40 mL vials
Soil – 4 oz. glass container with Teflon-lined lid

Preservative: Refrigerate or store on ice. Do not allow to freeze.

Handling: **Recommend next day receipt at lab.** Sample extraction must begin within 14 days of collection, then sample analysis within 40 days after extraction.

Method Technology: GC/FID Detection

Preparation Method: Extraction technique (liquid-liquid, waste dilution); Water- EPA 3510; Soil- EPA 3580

[Back to Organic Analytical Services](#)

Organic Chemistry


Name: RCRA Toxic Characteristic Leaching Procedure (TCLP)

Method: TCLP-Organics (VOCs, SVOCs, Pesticides, Herbicides)

NOTE: This test is NOT performed at UPHL and will be sent to a subcontracted laboratory for testing.

Groupings: RCRA Organic Testing

Application: Wastewater, Soil, Oil



Herbicides

2,4-D
2,4,5-TP (Silvex)

Pesticides

Chlordane
Endrin
Heptachlor
Heptachlor Epoxide
Lindane
Methoxychlor
Toxaphene

VOCs

D029 1,1-Dichloroethene
D028 1,2-Dichloroethane
D027 1,4-Dichlorobenzene
D018 Benzene
D019 Carbon Tetrachloride
D021 Chlorobenzene
D022 Chloroform
D035 Methyl Ethyl Ketone
D039 Tetrachloroethene
D040 Trichloroethene
D043 Vinyl Chloride

SVOCs

D030 2,4-Dinitrotoluene
D041 2,4,5-Trichlorophenol
D042 2,4,6-Trichlorophenol
D032 Hexachlorobenzene
D033 Hexachlorobutadiene
D034 Hexachloroethane
D024 m-Cresol
D036 Nitrobenzene
D023 o-Cresol
D025 p-Cresol
D037 Pentachlorophenol
D038 Pyridine

Instructions for Collection: Slowly fill to top of container.

Required Liquids – 4 L amber glass bottle

Containers/Volume: **Soil & Oil** – 4 oz. amber glass container with Teflon-lined lid

Preservative: Refrigerate or store on ice. Do not allow to freeze.

Handling: **Recommend next day receipt at lab.** Prep and analysis times:

SVOCs: 7 days to TCLP/40 days to analyze.

VOCs: 14 days to TCLP/40 days to analyze.

Pesticides: 7 days to extract/40 days to analyze.

Solids: 14 days to extract/ 40 days to analyze.

Herbicides: 14 days to extract/28 days to analyze.

Method Technology: GCMS Detection

Preparation Method: Leaching Procedure & Extraction Technique

[Back to Organic Analytical Services](#)

Organic Chemistry

Name: Volatile Organic Compounds (VOCs)

Methods: EPA 524.2, EPA 624, EPA 8260 /Subcontracted

NOTE: This test is NOT performed at UPHL and will be sent to a subcontracted laboratory for testing.

Application: Drinking Water, Surface Water, Wastewater, Groundwater, Solids



Analytes: Refer to Program Specific Indexes

CWA: [Page 67](#)

RCRA: [Page 72](#)

SDWA: [Page 77](#)

Note: If using Encore sampler, please bring to lab as soon as possible (within 48 hours).

Instructions for Collection: **Caution:** For surface waters and groundwater samples, check to see if HCl preservative reacts with source water (foams, effervesces, etc.). If a reaction occurs, do not add HCl. For aqueous samples, fill vials with sample source to top of container. Add 2 drops HCl acid to each vial, more drops needed if highly buffered source. Be sure not to over fill to prevent loss of preservative. Check for air bubbles by inverting, and fill the remaining portion if bubbles are found. **There must be no headspace.**

Required Containers/Volume: **Drinking Water** – 3/40mL vials, plus 1 trip blank
Surface Water – 4/40 mL, plus 1 trip blank
Groundwater – 4/40 mL vials, plus 1 trip blank, prepared and delivered with sample bottles.

Note: For chlorinated sites, use specially prepared 4/40 mL vials.

Soil, solids, sludge – 4 oz. glass container with Teflon-lined lid

Preservative: **Drinking Water** – 25 mg ascorbic acid, HCl to pH <2.
Surface Water – 10 mg Na₂S₂O₃ for chlorinated sites and HCl if needed.
Groundwater – 10 mg Na₂S₂O₃ for chlorinated sites and HCl if needed.
 All sample types, refrigerate or store on ice. Do not allow to freeze, 4-6°C.

Handling: **Recommend next day receipt at lab. For water, sample analysis must begin within 14 days after collection. For Soil, extraction must occur within 14 days after collection and then 14 days for analysis after extraction.**

Method Technology: Purge and trap technique followed by GCMS detection

Preparation Method: Water (EPA 624 and EPA 8260): EPA 3050. Soil (EPA 8260): EPA 3585

[Back to Organic Analytical Services](#)


Environmental Water Microbiology

Name: Heterotrophic Plate Count (HPC)

Method: SM 9215B

Application: Drinking Water, Pool/Spa Water, Reagent Water (deionized, distilled, etc.), Surface Water

Analytes: Heterotrophic Plate Count (Total bacteria count)



Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to top of container, be sure not to over fill to prevent loss of preservative.

Required Containers/Volume: Sterile plastic bottle; 120 mL Water Bacteriology

Preservative: 10 mg $\text{Na}_2\text{S}_2\text{O}_3$ (sodium thiosulfate)

Handling: Refrigerate or store on ice. Do not allow sample to freeze. **Recommend same day receipt at lab.** Sample analysis must begin within 8 hours of time of collection.

Method Technology: Pour Plate Agar

[Back to Env Water Micro Analytical Services](#)

Environmental Water Microbiology

Name: Legionella
Method: SM 9260J (CDC ELITE)

Application: Drinking Water (hot water tanks),
Air Handling (swamp coolers,
evaporators, etc.), cotton swabs

Analytes: *Legionella pneumophila*
(serogroups 1-14);
select *Legionella* species

The Utah Public Health Laboratory is a member of
the CDC's ELITE program for *Legionella* testing.



Instructions for Collection: Allow source to flow for a few minutes until water
temperature stabilizes. Slowly fill to top of container.

Required Containers/Volume: Unpreserved plastic bottle; 500 mL – 2 L.

Preservative: May be preserved with sodium thiosulfate if water
sample is chlorinated.

Handling: Recommend same day or next day receipt at lab.
Testing must begin within 48 hours of receiving sample.

Method Technology: Plate culture w/ Latex-Agglutination confirmation


The Utah Public Health Laboratory is a member
of the CDC's ELITE program for *Legionella* testing.

[Back to Env Water Micro Analytical Services](#)

Environmental Water Microbiology

Name: Total Coliforms/*E.coli*.
Method: SM 9223B (Quanti-Tray)

Application: Drinking Water, Surface Water



Analytes: Total Coliforms (Presence/Absence; MPN)
E.coli. (Presence/Absence; MPN)

Instructions for Collection: Allow source to flow for a few minutes until water temperature stabilizes. Slowly fill to above the 100 mL line but below the 120 mL line. Do not over fill to prevent loss of preservative.

Required Containers/Volume: Sterile Plastic Bottle; 120 mL Water Bacteriology

Preservative: 10 mg $\text{Na}_2\text{S}_2\text{O}_3$ (sodium thiosulfate)

Handling: Refrigerate or store on ice. Do not allow sample to freeze. **Recommend same day receipt at lab.** Sample analysis must begin within 30 hours after collection, except for Surface Water which must begin within 8 hours.

Method Technology: Chromofluorogenic; Quanti-Tray w/ 18-Hr Collilert, 24-Hr Collilert, or Colisure Reagents.

[Back to Env Water Micro Analytical Services](#)

Program Specific Indexes

Clean Water Act (CWA)

Analyte	Method	UPHL Section	Page #
Alkalinity	SM 2320B	Inorganic Chemistry	13
Aluminum (Al)	EPA 200.8	Metals	47
Ammonia (NH ₃)	EPA 350.1	Inorganic Chemistry	14
Antimony (Sb)	EPA 200.8	Metals	47
Arsenic (As)	EPA 200.8	Metals	39
Barium (Ba)	EPA 200.8	Metals	47
Beryllium (Be)	EPA 200.8	Metals	47
BOD	SM 5210B	<i>Subcontracted</i>	15
Boron (B)	EPA 200.7	Metals	47
Cadmium (Cd)	EPA 200.8	Metals	47
Calcium (Ca)	EPA 200.7	Metals	47
CBOD	SM 5210B	<i>Subcontracted</i>	15
Chloride	EPA 325.2	Inorganic Chemistry	17
Chlorophyll-A	SM 10200H	Inorganic Chemistry	48
Chromium (Cr)	EPA 200.8	Metals	47
Chromium-VI	EPA 218.7	<i>Subcontracted</i>	41
Cobalt (Co)	EPA 200.8	Metals	47
Conductivity	SM 2510B	Inorganic Chemistry	19
Copper (Cu)	EPA 200.8	Metals	47
Cyanide	EPA 335.4	Inorganic Chemistry	21
<i>E. coli</i>	SM 9223B	Environmental Microbiology	68
Fluoride	EPA 300.0	Inorganic Chemistry	22
Hardness as CaCO ₃	SM 2340B	Inorganic Chemistry	44
Heterotrophic Plate Count	SM 9215B	Environmental Microbiology	66
Iron (Fe)	EPA 200.7	Metals	47
Lead (Pb)	EPA 200.8	Metals	47
Magnesium (Mg)	EPA 200.7	Metals	47
Manganese (Mn)	EPA 200.8	Metals	47
Mercury (Hg)	EPA 200.8; EPA 245.1	Metals	40
Molybdenum (Mo)	EPA 200.8	Metals	47
Nickel (Ni)	EPA 200.8	Metals	47
Nitrate+Nitrite	EPA 353.2	Inorganic Chemistry	23
Nitrite (only)	EPA 353.2	Inorganic Chemistry	23
Oil and Grease	EPA 1664	<i>Subcontracted</i>	59
Phosphate	EPA 365.1	Inorganic Chemistry-Nutrients	28
pH	EPA 150.1	Inorganic Chemistry	27
Potassium (K)	EPA 200.7	Metals	47
Selenium (Se)	EPA 200.8	Metals	47
Silica	EPA 370.1	Inorganic Chemistry	29
Silver (Ag)	EPA 200.8	Metals	47

Program Specific Indexes

Clean Water Act (CWA)

Analyte	Method	UPHL Section	Page #
Sodium (Na)	EPA 200.7	Metals	47
Sulfate	EPA 375.2	Inorganic Chemistry	30
Sulfide	EPA 376.2	Inorganic Chemistry	31
Thallium (Tl)	EPA 200.8	Metals	47
Total Organic Carbon	SM 5310B	Inorganic Chemistry	51
Total Coliforms	SM 9223B	Environmental Microbiology	68
Total Dissolved Solids	SM 2540C	Inorganic Chemistry	32
Total Suspended Solids	SM 2540D	Inorganic Chemistry	32
Turbidity	EPA 180.1	Inorganic Chemistry	33
UV254	SM 5910B-UV254	Inorganic Chemistry	52
Vanadium	EPA 200.8	Metals	47
Zinc	EPA 200.8	Metals	47
4,4'-DDD	EPA 608	<i>Subcontracted</i>	61
4,4'-DDE	EPA 608	<i>Subcontracted</i>	61
4,4'-DDT	EPA 608	<i>Subcontracted</i>	61
Aldrin	EPA 608	<i>Subcontracted</i>	61
alpha-BHC	EPA 608	<i>Subcontracted</i>	61
beta-BHC	EPA 608	<i>Subcontracted</i>	61
Chlordane	EPA 608	<i>Subcontracted</i>	61
delta-BHC	EPA 608	<i>Subcontracted</i>	61
Dieldrin	EPA 608	<i>Subcontracted</i>	61
Endosulfan I	EPA 608	<i>Subcontracted</i>	61
Endosulfan II	EPA 608	<i>Subcontracted</i>	61
Endosulfan sulfate	EPA 608	<i>Subcontracted</i>	61
Endrin	EPA 608	<i>Subcontracted</i>	61
Endrin aldehyde	EPA 608	<i>Subcontracted</i>	61
gamma-BHC	EPA 608	<i>Subcontracted</i>	61
Heptachlor	EPA 608	<i>Subcontracted</i>	61
Heptachlor epoxide	EPA 608	<i>Subcontracted</i>	61
Methoxychlor	EPA 608	<i>Subcontracted</i>	61
PCB-1016	EPA 608	<i>Subcontracted</i>	61
PCB-1221	EPA 608	<i>Subcontracted</i>	61
PCB-1232	EPA 608	<i>Subcontracted</i>	61
PCB-1242	EPA 608	<i>Subcontracted</i>	61
PCB-1248	EPA 608	<i>Subcontracted</i>	61
PCB-1254	EPA 608	<i>Subcontracted</i>	61
PCB-1260	EPA 608	<i>Subcontracted</i>	61
Toxaphene	EPA 608	<i>Subcontracted</i>	61
1,1-Dichloroethane	EPA 624	<i>Subcontracted</i>	65
1,1-Dichloroethene	EPA 624	<i>Subcontracted</i>	65
1,1-Dichloropropene	EPA 624	<i>Subcontracted</i>	65

Program Specific Indexes

Clean Water Act (CWA)

Analyte	Method	UPHL Section	Page #
1,1,1-Trichloroethane	EPA 624	<i>Subcontracted</i>	65
1,1,1,2- Tetrachloroethane	EPA 624	<i>Subcontracted</i>	65
1,1,2-Trichloroethane	EPA 624	<i>Subcontracted</i>	65
1,1,2-Trichloro-1,2,2-Trifluoroethane (Freon 113)	EPA 624	<i>Subcontracted</i>	65
1,1,2,2-Tetrachloroethane	EPA 624	<i>Subcontracted</i>	65
1,2-Dibromo-3-chloropropane	EPA 624	<i>Subcontracted</i>	65
1,2-Dichlorobenzene	EPA 624	<i>Subcontracted</i>	65
1,2-Dichloroethane	EPA 624	<i>Subcontracted</i>	65
1,2-Dichloropropane	EPA 624	<i>Subcontracted</i>	65
1,2-Dichlorotoluene	EPA 624	<i>Subcontracted</i>	65
1,2,3-Trichlorobenzene	EPA 624	<i>Subcontracted</i>	65
1,2,3-Trichloropropane	EPA 624	<i>Subcontracted</i>	65
1,2,4-Trichlorobenzene	EPA 624	<i>Subcontracted</i>	65
1,2,4-Trimethylbenzene	EPA 624	<i>Subcontracted</i>	65
1,3-Dichlorobenzene	EPA 624	<i>Subcontracted</i>	65
1,3-Dichloropropane	EPA 624	<i>Subcontracted</i>	65
1,3,5-Trimethylbenzene	EPA 624	<i>Subcontracted</i>	65
1,4-Dichlorobenzene	EPA 624	<i>Subcontracted</i>	65
1,4-Dichlorotoluene	EPA 624	<i>Subcontracted</i>	65
1,4-Isopropyltoluene	EPA 624	<i>Subcontracted</i>	65
2,2-Dichloropropane	EPA 624	<i>Subcontracted</i>	65
Benzene	EPA 624	<i>Subcontracted</i>	65
Bromobenzene	EPA 624	<i>Subcontracted</i>	65
Bromochloromethane	EPA 624	<i>Subcontracted</i>	65
Bromodichloromethane	EPA 624	<i>Subcontracted</i>	65
Bromoform	EPA 624	<i>Subcontracted</i>	65
Bromomethane	EPA 624	<i>Subcontracted</i>	65
Carbon tetrachloride	EPA 624	<i>Subcontracted</i>	65
Chlorobenzene	EPA 624	<i>Subcontracted</i>	65
Chlorodibromomethane	EPA 624	<i>Subcontracted</i>	65
Chloroethane	EPA 624	<i>Subcontracted</i>	65
Chloroform	EPA 624	<i>Subcontracted</i>	65
Chloromethane	EPA 624	<i>Subcontracted</i>	65
cis-1,2-Dichloroethene	EPA 624	<i>Subcontracted</i>	65
cis-1,3-Dichloropropene	EPA 624	<i>Subcontracted</i>	65
Dibromomethane	EPA 624	<i>Subcontracted</i>	65
Dichlorodifluoromethane	EPA 624	<i>Subcontracted</i>	65
Ethylbenzene	EPA 624	<i>Subcontracted</i>	65
Ethylene dibromide	EPA 624	<i>Subcontracted</i>	65
Hexachlorobutadiene	EPA 624	<i>Subcontracted</i>	65
Isopropylbenzene	EPA 624	<i>Subcontracted</i>	65
Methylene chloride	EPA 624	<i>Subcontracted</i>	65

Program Specific Indexes

Clean Water Act (CWA)

Analyte	Method	UPHL Section	Page #
MTBE	EPA 624	<i>Subcontracted</i>	65
Napthalene	EPA 624	<i>Subcontracted</i>	65
n-Butylbenzene	EPA 624	<i>Subcontracted</i>	65
n-Propylbenzene	EPA 624	<i>Subcontracted</i>	65
sec-Butylbenzene	EPA 624	<i>Subcontracted</i>	65
Styrene	EPA 624	<i>Subcontracted</i>	65
tert-Butylbenzene	EPA 624	<i>Subcontracted</i>	65
Tetrachloroethene (PCE)	EPA 624	<i>Subcontracted</i>	65
Toluene	EPA 624	<i>Subcontracted</i>	65
trans-1,2-Dichloroethene	EPA 624	<i>Subcontracted</i>	65
trans-1,3-Dichloropropene	EPA 624	<i>Subcontracted</i>	65
Trichloroethene (TCE)	EPA 624	<i>Subcontracted</i>	65
Trichlorofluoromethane	EPA 624	<i>Subcontracted</i>	65
Vinyl chloride	EPA 624	<i>Subcontracted</i>	65
Xylene	EPA 624	<i>Subcontracted</i>	65
1,2-Dichlorobenzene	EPA 625	<i>Subcontracted</i>	60
1,2,4-Trichlorobenzene	EPA 625	<i>Subcontracted</i>	60
1,3-Dichlorobenzene	EPA 625	<i>Subcontracted</i>	60
1,4-Dichlorobenzene	EPA 625	<i>Subcontracted</i>	60
2-Chloronaphthalene	EPA 625	<i>Subcontracted</i>	60
2-Chlorophenol	EPA 625	<i>Subcontracted</i>	60
2-Methyl naphthalene	EPA 625	<i>Subcontracted</i>	60
2-Methyl phenol	EPA 625	<i>Subcontracted</i>	60
2-Methyl-4,6-dinitrophenol	EPA 625	<i>Subcontracted</i>	60
2-Nitroaniline	EPA 625	<i>Subcontracted</i>	60
2-Nitrophenol	EPA 625	<i>Subcontracted</i>	60
2,4-Dichlorophenol	EPA 625	<i>Subcontracted</i>	60
2,4-Dimethylphenol	EPA 625	<i>Subcontracted</i>	60
2,4-Dinitrophenol	EPA 625	<i>Subcontracted</i>	60
2,4-Dinitrotoluene	EPA 625	<i>Subcontracted</i>	60
2,4,6-Trichlorophenol	EPA 625	<i>Subcontracted</i>	60
2,6-Dinitrotoluene	EPA 625	<i>Subcontracted</i>	60
3-Methyl phenol	EPA 625	<i>Subcontracted</i>	60
3-Nitroaniline	EPA 625	<i>Subcontracted</i>	60
3,3'-Dichlorobenzidine	EPA 625	<i>Subcontracted</i>	60
4-Bromophenyl phenyl ether	EPA 625	<i>Subcontracted</i>	60
4-Chloroaniline	EPA 625	<i>Subcontracted</i>	60
4-Chlorophenyl phenyl ether	EPA 625	<i>Subcontracted</i>	60
4-Chloro-3-methyl phenol	EPA 625	<i>Subcontracted</i>	60
4-Methyl phenol	EPA 625	<i>Subcontracted</i>	60
4-Nitroaniline	EPA 625	<i>Subcontracted</i>	60
4-Nitrophenol	EPA 625	<i>Subcontracted</i>	60

Program Specific Indexes

Clean Water Act (CWA)

Analyte	Method	UPHL Section	Page #
Acenaphthene	EPA 625	Subcontracted	60
Acenaphthylene	EPA 625	Subcontracted	60
Aniline	EPA 625	Subcontracted	60
Anthracene	EPA 625	Subcontracted	60
Benzidine	EPA 625	Subcontracted	60
Benzo (a) anthracene	EPA 625	Subcontracted	60
Benzo (a) pyrene	EPA 625	Subcontracted	60
Benzo (b) fluoranthene	EPA 625	Subcontracted	60
Benzo (g,h,i) perylene	EPA 625	Subcontracted	60
Benzo (k) fluoranthene	EPA 625	Subcontracted	60
Benzylbutylphthalate	EPA 625	Subcontracted	60
Benzyl alcohol	EPA 625	Subcontracted	60
Benzoic acid	EPA 625	Subcontracted	60
bis (2-chloroethyl) ether	EPA 625	Subcontracted	60
bis (2-chloroethoxy) methane	EPA 625	Subcontracted	60
bis (2-chloroisopropyl) ether	EPA 625	Subcontracted	60
bis (2-ethylhexyl) phthalate	EPA 625	Subcontracted	60
Chrysene	EPA 625	Subcontracted	60
Dibenzo (a,h) anthracene	EPA 625	Subcontracted	60
Dibenzofuran	EPA 625	Subcontracted	60
Diethyl phthalate	EPA 625	Subcontracted	60
Dimethyl phthalate	EPA 625	Subcontracted	60
Di-n-butyl phthalate	EPA 625	Subcontracted	60
Di-n-octyl phthalate	EPA 625	Subcontracted	60
Fluoranthene	EPA 625	Subcontracted	60
Fluorene	EPA 625	Subcontracted	60
Hexachlorobenzene	EPA 625	Subcontracted	60
Hexachlorobutadiene	EPA 625	Subcontracted	60
Hexachlorocyclopentadiene	EPA 625	Subcontracted	60
Hexachloroethane	EPA 625	Subcontracted	60
Indeno (1,2,3-cd) pyrene	EPA 625	Subcontracted	60
Isophorone	EPA 625	Subcontracted	60
n-Nitrosodimethylamine	EPA 625	Subcontracted	60
n-Nitrosodiphenylamine	EPA 625	Subcontracted	60
n-Nitrosodipropylamine	EPA 625	Subcontracted	60
Naphthalene	EPA 625	Subcontracted	60
Nitrobenzene	EPA 625	Subcontracted	60
Pentachlorophenol	EPA 625	Subcontracted	60
Phenanthrene	EPA 625	Subcontracted	60
Phenol	EPA 625	Subcontracted	60
Pyrene	EPA 625	Subcontracted	60

Program Specific Indexes

Resource Conservation & Recovery Act (RCRA)

Analyte	Method	EPA HW#	UPHL Section	Page #
4,4'-DDD	EPA 8081	U060	<i>Subcontracted</i>	61
4,4'-DDE	EPA 8081		<i>Subcontracted</i>	61
4,4'-DDT	EPA 8081	U061	<i>Subcontracted</i>	61
Aldrin	EPA 8081	P005	<i>Subcontracted</i>	61
alpha-BHC	EPA 8081	U129	<i>Subcontracted</i>	61
beta-BHC	EPA 8081		<i>Subcontracted</i>	61
Chlordane	EPA 8081	D020	<i>Subcontracted</i>	61
delta-BHC	EPA 8081		<i>Subcontracted</i>	61
Dieldrin	EPA 8081	P037	<i>Subcontracted</i>	61
Endosulfan I	EPA 8081	P050	<i>Subcontracted</i>	61
Endosulfan II	EPA 8081	P050	<i>Subcontracted</i>	61
Endosulfan sulfate	EPA 8081	P050	<i>Subcontracted</i>	61
Endrin	EPA 8081	D012	<i>Subcontracted</i>	61
Endrin aldehyde	EPA 8081	D012	<i>Subcontracted</i>	61
gamma-BHC (Lindane)	EPA 8081	D013	<i>Subcontracted</i>	61
Heptachlor	EPA 8081	D031	<i>Subcontracted</i>	61
Heptachlor epoxide	EPA 8081	D031	<i>Subcontracted</i>	61
Methoxychlor	EPA 8081	D014	<i>Subcontracted</i>	61
PCB-1016	EPA 8082		<i>Subcontracted</i>	61
PCB-1221	EPA 8082		<i>Subcontracted</i>	61
PCB-1232	EPA 8082		<i>Subcontracted</i>	61
PCB-1242	EPA 8082		<i>Subcontracted</i>	61
PCB-1248	EPA 8082		<i>Subcontracted</i>	61
PCB-1254	EPA 8082		<i>Subcontracted</i>	61
PCB-1260	EPA 8082		<i>Subcontracted</i>	61
Toxaphene	EPA 8081	D015	<i>Subcontracted</i>	61
*Add-ons	EPA 8081		<i>Subcontracted</i>	61
2,4-D	EPA 8151	D016	<i>Subcontracted</i>	58
2,4,5-TP (Silvex)	EPA 8151	D017	<i>Subcontracted</i>	58
Dalapon	EPA 8151		<i>Subcontracted</i>	58
Dinoseb	EPA 8151	P020	<i>Subcontracted</i>	58
Pentachlorophenol	EPA 8151	D037	<i>Subcontracted</i>	58
Picloram	EPA 8151		<i>Subcontracted</i>	58
*Add-ons	EPA 8151		<i>Subcontracted</i>	58
1,1-Dichloroethane	EPA 8260		<i>Subcontracted</i>	65
1,1-Dichloroethene	EPA 8260	D029	<i>Subcontracted</i>	65
1,1-Dichloropropene	EPA 8260		<i>Subcontracted</i>	65
1,1,1-Trichloroethane	EPA 8260		<i>Subcontracted</i>	65
1,1,1,2-Tetrachloroethane	EPA 8260	U208	<i>Subcontracted</i>	65
1,1,2-Trichloroethane	EPA 8260	U227	<i>Subcontracted</i>	65

*Tentative analysis by special request, record specific analyte on test request form and schedule at 801-965-2400

Program Specific Indexes

Resource Conservation & Recovery Act (RCRA)

Analyte	Method	EPA HW#	UPHL Unit	Page #
1,1,2-Trichloro-1,2,2-trifluoroethane (Freon 113)	EPA 8260		Subcontracted	65
1,1,2,2-Tetrachloroethane	EPA 8260	U209	Subcontracted	65
1,2-Dibromo-3-chloropropane	EPA 8260	U066	Subcontracted	65
1,2-Dichlorobenzene	EPA 8260	U070	Subcontracted	65
1,2-Dichloroethane	EPA 8260	D028	Subcontracted	65
1,2-Dichloropropane	EPA 8260	U083	Subcontracted	65
1,2-Dichlorotoluene	EPA 8260		Subcontracted	65
1,2,3-Trichlorobenzene	EPA 8260		Subcontracted	65
1,2,3-Trichloropropane	EPA 8260		Subcontracted	65
1,2,4-Trichlorobenzene	EPA 8260		Subcontracted	65
1,2,4-Trimethylbenzene	EPA 8260		Subcontracted	65
1,3-Dichlorobenzene	EPA 8260	U071	Subcontracted	65
1,3-Dichloropropane	EPA 8260	U084	Subcontracted	65
1,3,5-Trimethylbenzene	EPA 8260		Subcontracted	65
1,4-Dichlorobenzene	EPA 8260	D027	Subcontracted	65
1,4-Dichlorotoluene	EPA 8260	U072	Subcontracted	65
1,4-Isopropyltoluene	EPA 8260		Subcontracted	65
2,2-Dichloropropane	EPA 8260		Subcontracted	65
Benzene	EPA 8260	D018	Subcontracted	65
Bromobenzene	EPA 8260		Subcontracted	65
Bromochloromethane	EPA 8260		Subcontracted	65
Bromodichloromethane	EPA 8260		Subcontracted	65
Bromoform	EPA 8260	U225	Subcontracted	65
Bromomethane	EPA 8260	U029	Subcontracted	65
Carbon tetrachloride	EPA 8260	D019	Subcontracted	65
Chloroethane	EPA 8260	U045	Subcontracted	65
Chloroform	EPA 8260	D022	Subcontracted	65
Chloromethane	EPA 8260	U045	Subcontracted	65
cis-1,2-Dichloroethene	EPA 8260		Subcontracted	65
cis-1,3-Dichloropropene	EPA 8260		Subcontracted	65
Dibromomethane	EPA 8260	U068	Subcontracted	65
Dichlorodifluoromethane	EPA 8260	U075	Subcontracted	65
Ethylbenzene	EPA 8260		Subcontracted	65
Ethylene dibromide	EPA 8260	U067	Subcontracted	65
Hexachlorobutadiene	EPA 8260	D033	Subcontracted	65
Isopropylbenzene	EPA 8260		Subcontracted	65
Methyl ethyl ketone*	EPA 8260	D035	Subcontracted	65
Methylene chloride	EPA 8260	U080	Subcontracted	65
MTBE	EPA 8260		Subcontracted	65
Napthalene	EPA 8260		Subcontracted	65

*Tentative analysis by special request, record specific analyte on test request form and schedule at 801-965-2400

Program Specific Indexes

Resource Conservation & Recovery Act (RCRA)

Analyte	Method	EPA HW#	UPHL Unit	Page #
n-Butylbenzene	EPA 8260		Subcontracted	65
n-Propylbenzene	EPA 8260		Subcontracted	65
sec-Butylbenzene	EPA 8260		Subcontracted	65
Styrene	EPA 8260		Subcontracted	65
tert-Butylbenzene	EPA 8260		Subcontracted	65
Tetrachloroethene (PCE)	EPA 8260	D039	Subcontracted	65
Toluene	EPA 8260	U220	Subcontracted	65
trans-1,2-Dichloroethene	EPA 8260	U079	Subcontracted	65
trans-1,3-Dichloropropene	EPA 8260	U084	Subcontracted	65
Trichloroethene (TCE)	EPA 8260	D040	Subcontracted	65
Trichlorofluoromethane	EPA 8260	U121	Subcontracted	65
Vinyl chloride	EPA 8260	D043	Subcontracted	65
Xylene	EPA 8260	U239	Subcontracted	65
*Add-ons	EPA 8260		Subcontracted	65
1,2-Dichlorobenzene	EPA 8270		Subcontracted	60
1,2,4-Trichlorobenzene	EPA 8270		Subcontracted	60
1,3-Dichlorobenzene	EPA 8270		Subcontracted	60
1,4-Dichlorobenzene	EPA 8270		Subcontracted	60
2-Chloronaphthalene	EPA 8270	U047	Subcontracted	60
2-Chlorophenol	EPA 8270	U048	Subcontracted	60
2-Methyl naphthalene	EPA 8270		Subcontracted	60
2-Methyl phenol (o-Cresol)	EPA 8270	D023	Subcontracted	60
2-Methyl-4,6-dinitrophenol	EPA 8270		Subcontracted	60
2-Nitroaniline	EPA 8270		Subcontracted	60
2-Nitrophenol	EPA 8270		Subcontracted	60
2,4-Dichlorophenol	EPA 8270	U081	Subcontracted	60
2,4-Dimethylphenol	EPA 8270	U101	Subcontracted	60
2,4-Dinitrophenol	EPA 8270	P048	Subcontracted	60
2,4-Dinitrotoluene	EPA 8270	D030	Subcontracted	60
2,4,5-Trichlorophenol	EPA 8270	D041	Subcontracted	60
2,4,6-Trichlorophenol	EPA 8270	D042	Subcontracted	60
2,6-Dinitrotoluene	EPA 8270	U106	Subcontracted	60
3-Methyl phenol (m-Cresol)	EPA 8270	D024	Subcontracted	60
3-Nitroaniline	EPA 8270		Subcontracted	60
3,3'-Dichlorobenzidine	EPA 8270	U073	Subcontracted	60
4-Bromophenyl phenyl ether	EPA 8270	U030	Subcontracted	60
4-Chloroaniline	EPA 8270	P024	Subcontracted	60
4-Chlorophenyl phenyl ether	EPA 8270		Subcontracted	60
4-Chloro-3-methyl phenol	EPA 8270	U039	Subcontracted	60
4-Methyl phenol (p-Cresol)	EPA 8270	D025	Subcontracted	60

*Tentative analysis by special request, record specific analyte on test request form and schedule at 801-965-2400

Program Specific Indexes

Resource Conservation & Recovery Act (RCRA)

Analyte	Method	EPA HW#	UPHL Unit	Page #
4-Nitroaniline	EPA 8270	P077	Subcontracted	60
4-Nitrophenol	EPA 8270	U170	Subcontracted	60
Acenaphthene	EPA 8270		Subcontracted	60
Acenaphthylene	EPA 8270		Subcontracted	60
Aniline	EPA 8270	U012	Subcontracted	60
Anthracene	EPA 8270		Subcontracted	60
Benzidine	EPA 8270	U021	Subcontracted	60
Benzo (a) anthracene	EPA 8270	U018	Subcontracted	60
Benzo (a) pyrene	EPA 8270	U022	Subcontracted	60
Benzo (b) fluoranthene	EPA 8270		Subcontracted	60
Benzo (g,h,i) perylene	EPA 8270		Subcontracted	60
Benzo (k) fluoranthene	EPA 8270		Subcontracted	60
Benzylbutylphthalate	EPA 8270		Subcontracted	60
Benzyl alcohol	EPA 8270		Subcontracted	60
Benzoic acid	EPA 8270		Subcontracted	60
bis (2-chloroethyl) ether	EPA 8270	U025	Subcontracted	60
bis (2-chloroethoxy) methane	EPA 8270	U046	Subcontracted	60
bis (2-chloroisopropyl) ether	EPA 8270	U027	Subcontracted	60
bis (2-ethylhexyl) phthalate	EPA 8270	U028	Subcontracted	60
Chrysene	EPA 8270	U050	Subcontracted	60
Dibenzo (a,h) anthracene	EPA 8270	U063	Subcontracted	60
Dibenzofuran	EPA 8270		Subcontracted	60
Diethyl phthalate	EPA 8270	U088	Subcontracted	60
Dimethyl phthalate	EPA 8270	U102	Subcontracted	60
Di-n-butyl phthalate	EPA 8270	U069	Subcontracted	60
Di-n-octyl phthalate	EPA 8270	U107	Subcontracted	60
Fluoranthene	EPA 8270	U120	Subcontracted	60
Fluorene	EPA 8270		Subcontracted	60
Hexachlorobenzene	EPA 8270	D032	Subcontracted	60
Hexachlorobutadiene	EPA 8270	D033	Subcontracted	60
Hexachlorocyclopentadiene	EPA 8270	U130	Subcontracted	60
Hexachloroethane	EPA 8270	D034	Subcontracted	60
Ideno (1,2,3-cd) pyrene	EPA 8270	U137	Subcontracted	60
Isophorone	EPA 8270		Subcontracted	60
n-Nitrosodimethylamine	EPA 8270	P082	Subcontracted	60
n-Nitrosodiphenylamine	EPA 8270		Subcontracted	60
n-Nitrosodipropylamine	EPA 8270		Subcontracted	60
Naphthalene	EPA 8270	U165	Subcontracted	60
Nitrobenzene	EPA 8270	D036	Subcontracted	60
Pentachlorophenol	EPA 8270	D037	Subcontracted	60
Phenanthrene	EPA 8270		Subcontracted	60

Program Specific Indexes

Resource Conservation & Recovery Act (RCRA)

Analyte	Method	EPA HW#	UPHL Unit	Page #
Phenol	EPA 8270	U188	<i>Subcontracted</i>	60
Pyrene	EPA 8270		<i>Subcontracted</i>	60
*Pyridine	EPA 8270	D038	<i>Subcontracted</i>	60
*Add-ons	EPA 8270		<i>Subcontracted</i>	60

*Tentative analysis by special request, record specific analyte on test request form and schedule at 801-965-2400

Program Specific Indexes

Safe Drinking Water Act (SDWA)

Analyte	Method	UPHL Unit	Page #
2,4-D	EPA 515.1	<i>Subcontracted</i>	58
2,4,5-TP (Silvex)	EPA 515.1	<i>Subcontracted</i>	58
3-Hydroxycarbofuran	EPA 531.1	<i>Subcontracted</i>	56
Aldicarb (Temik)	EPA 531.1	<i>Subcontracted</i>	56
Aldicarb sulfone	EPA 531.1	<i>Subcontracted</i>	56
Aldicarb sulfoxide	EPA 531.1	<i>Subcontracted</i>	56
Alkalinity	SM 2320B	Inorganic Chemistry	13
Aluminum	EPA 200.8	Metals	47
Ammonia	SM 4500 NH3 H	Inorganic Chemistry	14
Antimony	EPA 200.8	Metals	47
Arsenic	EPA 200.8	Metals	39
Barium	EPA 200.8	Metals	47
Beryllium	EPA 200.8	Metals	47
Boron	EPA 200.7	Metals	47
Bromate	EPA 300.1	Inorganic Chemistry	16
Bromide	EPA 300.1	Inorganic Chemistry	16,17
Bromodichloromethane	EPA 524.2	Organic Chemistry	54
Bromoform	EPA 524.2	Organic Chemistry	54
Cadmium	EPA 200.8	Metals	47
Calcium	EPA 200.7	Metals	47
Carbaryl (Sevin)	EPA 531.1	<i>Subcontracted</i>	56
Carbofuran (Furadan)	EPA 531.1	<i>Subcontracted</i>	56
Chlorate	EPA 300.1	Inorganic Chemistry	16
Chloride	EPA 300.0	Inorganic Chemistry	17
Chlorite	EPA 300.1	Inorganic Chemistry	16
Chlorodibromomethane	EPA 524.2	Organic Chemistry	54
Chloroform	EPA 524.2	Organic Chemistry	54
Chromium	EPA 200.8	Metals	47
Chromium-IV	EPA 218.7	<i>Subcontracted</i>	41
Cobalt	EPA 200.8	Metals	47
Color	SM 2120B	Inorganic Chemistry	18
Conductivity	EPA 120.1	Inorganic Chemistry	19
Copper	EPA 200.8	Metals	47
Corrosivity (Langelier Index)	SM 2330B	Inorganic Chemistry	20
Cyanide	EPA 335.4	Inorganic Chemistry	21
Dalapon	EPA 515.1	<i>Subcontracted</i>	58
Dibromoacetic acid	SM 6251B	Organic Chemistry	50
Dichloroacetic acid	SM 6251B	Organic Chemistry	50
Dinoseb	EPA 515.1	<i>Subcontracted</i>	58
<i>E. coli</i>	SM 9223B (Quanti-Tray)	Environmental Microbiology	68

Program Specific Indexes

Safe Drinking Water Act (SDWA)

Analyte	Method	UPHL Unit	Page #
Fluoride	EPA 300.0	Inorganic Chemistry	22
Hardness	SM 2340B	Inorganic Chemistry/Metals	44
Heterotrophic Plate Count	SM 9215B	Environmental Microbiology	66
Iron	EPA 200.7	Metals	47
Lead	EPA 200.8	Metals	47
<i>Legionella</i>	SM 9260J (CDC ELITE)	Environmental Microbiology	67
Magnesium	EPA 200.7	Metals	47
Manganese	EPA 200.8	Metals	47
Mercury	EPA 245.1	Metals	40
Methomyl	EPA 531.1	<i>Subcontracted</i>	56
Molybdenum	EPA 200.8	Metals	47
Monobromoacetic acid	SM 6215B	Organic Chemistry	50
Monochloroacetic acid	SM 6215B	Organic Chemistry	50
Nickel	EPA 200.8	Metals	47
Nitrate+Nitrite	EPA 353.2-NO2+NO3	Inorganic Chemistry	23
Nitrite (only)	EPA 353.2-NO2	Inorganic Chemistry	23
Odor	EPA 140.1	Inorganic Chemistry	25
Oxamyl (Vydate)	EPA 531.1	<i>Subcontracted</i>	56
Pentachlorophenol	EPA 515.1	<i>Subcontracted</i>	58
Perchlorate	EPA 314.0	<i>Subcontracted</i>	26
pH	EPA 150.1	Inorganic Chemistry	27
Phosphate	EPA 365.1	Inorganic Chemistry	28
Picloram	EPA 515.1	<i>Subcontracted</i>	58
Potassium	EPA 200.7	Metals	47
Selenium	EPA 200.8	Metals	47
Silica	EPA 370.1	Inorganic Chemistry	29
Silver	EPA 200.8	Metals	47
Sodium	EPA 200.7	Metals	47
Sulfate	EPA 300.0	Inorganic Chemistry	30
Sulfide	EPA 376.2	Inorganic Chemistry	31
Thallium	EPA 200.8	Metals	47
TOC	SM 5310B	Inorganic Chemistry	51
Total Coliform	SM 9223B (Quanti-Tray)	Environmental Microbiology	68
Total Dissolved Solids (TDS)	SM 2540C	Inorganic Chemistry	32
Total Suspended Solids(TSS)	SM 2540D	Inorganic Chemistry	32
Trichloroacetic acid	SM 6251B	Organic Chemistry	50
Turbidity	EPA 180.1	Inorganic Chemistry	33
UV254	SM 5910B	<i>Subcontracted</i>	52
Vanadium	EPA 200.8	Metals	47
Zinc	EPA 200.8	Metals	47
1,1-Dichloroethane	EPA 524.2	<i>Subcontracted</i>	65

Program Specific Indexes

Safe Drinking Water Act (SDWA)

Analyte	Method	UPHL Unit	Page #
1,1-Dichloroethene	EPA 524.2	Subcontracted	65
1,1-Dichloropropene	EPA 524.2	Subcontracted	65
1,1,1-Trichloroethane	EPA 524.2	Subcontracted	65
1,1,2-Trichloroethane	EPA 524.2	Subcontracted	65
1,1,1,2-Tetrachloroethane	EPA 524.2	Subcontracted	65
1,1,2,2-Tetrachloroethane	EPA 524.2	Subcontracted	65
1,2-Dibromo-3-chloropropane	EPA 524.2	Subcontracted	65
1,2-Dichlorobenzene	EPA 524.2	Subcontracted	65
1,2-Dichloroethane	EPA 524.2	Subcontracted	65
1,2-Dichloropropane	EPA 524.2	Subcontracted	65
1,2-Dichlorotoluene	EPA 524.2	Subcontracted	65
1,2,3-Trichlorobenzene	EPA 524.2	Subcontracted	65
1,2,3-Trichloropropane	EPA 524.2	Subcontracted	65
1,2,4-Trichlorobenzene	EPA 524.2	Subcontracted	65
1,2,4-Trimethylbenzene	EPA 524.2	Subcontracted	65
1,3-Dichlorobenzene	EPA 524.2	Subcontracted	65
1,3-Dichloropropane	EPA 524.2	Subcontracted	65
1,3,5-Trimethylbenzene	EPA 524.2	Subcontracted	65
1,4-Dichlorobenzene	EPA 524.2	Subcontracted	65
1,4-Dichlorotoluene	EPA 524.2	Subcontracted	65
1,4-Isopropyltoluene	EPA 524.2	Subcontracted	65
2,2-Dichloropropane	EPA 524.2	Subcontracted	65
Benzene	EPA 524.2	Subcontracted	65
Bromobenzene	EPA 524.2	Subcontracted	65
Bromochloromethane	EPA 524.2	Subcontracted	65
Bromodichloromethane	EPA 524.2	Subcontracted	65
Bromoform	EPA 524.2	Subcontracted	65
Bromomethane	EPA 524.2	Subcontracted	65
Carbon tetrachloride	EPA 524.2	Subcontracted	65
Chlorobenzene	EPA 524.2	Subcontracted	65
Chlorodibromomethane	EPA 524.2	Subcontracted	65
Chloroethane	EPA 524.2	Subcontracted	65
Chloroform	EPA 524.2	Subcontracted	65
Chloromethane	EPA 524.2	Subcontracted	65
cis-1,2-Dichloroethene	EPA 524.2	Subcontracted	65
cis-1,3-Dichloropropene	EPA 524.2	Subcontracted	65
Dibromomethane	EPA 524.2	Subcontracted	65
Dichlorodifluoromethane	EPA 524.2	Subcontracted	65
Ethylbenzene	EPA 524.2	Subcontracted	65
Ethylene dibromide	EPA 524.2	Subcontracted	65
Hexachlorobutadiene	EPA 524.2	Subcontracted	65

Program Specific Indexes

Safe Drinking Water Act (SDWA)

Analyte	Method	UPHL Unit	Page #
Isopropylbenzene	EPA 524.2	Subcontracted	65
Methylene chloride	EPA 524.2	Subcontracted	65
MTBE	EPA 524.2	Subcontracted	65
Napthalene	EPA 524.2	Subcontracted	65
n-Butylbenzene	EPA 524.2	Subcontracted	65
n-Propylbenzene	EPA 524.2	Subcontracted	65
sec-Butylbenzene	EPA 524.2	Subcontracted	65
Styrene	EPA 524.2	Subcontracted	65
tert-Butylbenzene	EPA 524.2	Subcontracted	65
Tetrachloroethene (PCE)	EPA 524.2	Subcontracted	65
Toluene	EPA 524.2	Subcontracted	65
trans-1,2-Dichloroethene	EPA 524.2	Subcontracted	65
trans-1,3-Dichloropropene	EPA 524.2	Subcontracted	65
Trichloroethene (TCE)	EPA 524.2	Subcontracted	65
Trichlorofluoromethane	EPA 524.2	Subcontracted	65
Vinyl chloride	EPA 524.2	Subcontracted	65
Xylene	EPA 524.2	Subcontracted	65
2,4-Dinitrotoluene	EPA 525.2	Subcontracted	60
2,6-Dinitrotoluene	EPA 525.2	Subcontracted	60
4,4'-DDE	EPA 525.2	Subcontracted	60
Acetochlor	EPA 525.2	Subcontracted	60
Alachlor	EPA 525.2	Subcontracted	60
Aldrin	EPA 525.2	Subcontracted	60
alpha-Chlordane	EPA 525.2	Subcontracted	60
Atrazine	EPA 525.2	Subcontracted	60
bis (2-ethylhexyl) adipate	EPA 525.2	Subcontracted	60
bis (2-ethylhexyl) phthalate	EPA 525.2	Subcontracted	60
Benzo (a) pyrene	EPA 525.2	Subcontracted	60
Bromocil	EPA 525.2	Subcontracted	60
Butachlor	EPA 525.2	Subcontracted	60
Chlorobiphenyl	EPA 525.2	Subcontracted	60
Cyanazine	EPA 525.2	Subcontracted	60
Dichlorobiphenyl	EPA 525.2	Subcontracted	60
Dieldrin	EPA 525.2	Subcontracted	60
Endrin	EPA 525.2	Subcontracted	60
EPTC	EPA 525.2	Subcontracted	60
gamma-Chlordane	EPA 525.2	Subcontracted	60
Heptachlor	EPA 525.2	Subcontracted	60
Heptachlor epoxide	EPA 525.2	Subcontracted	60
Heptachlorobiphenyl	EPA 525.2	Subcontracted	60
Hexachlorobenzene	EPA 525.2	Subcontracted	60

Program Specific Indexes

Safe Drinking Water Act (SDWA)

Analyte	Method	UPHL Unit	Page #
Hexachlorobiphenyl	EPA 525.2	<i>Subcontracted</i>	60
Hexachlorocyclopentadiene	EPA 525.2	<i>Subcontracted</i>	60
Lindane (gamma-BHC)	EPA 525.2	<i>Subcontracted</i>	60
Methoxychlor	EPA 525.2	<i>Subcontracted</i>	60
Metolachlor	EPA 525.2	<i>Subcontracted</i>	60
Metribuzin	EPA 525.2	<i>Subcontracted</i>	60
Molinate	EPA 525.2	<i>Subcontracted</i>	60
Octachlorobiphenyl	EPA 525.2	<i>Subcontracted</i>	60
Pentachlorobiphenyl	EPA 525.2	<i>Subcontracted</i>	60
Pentachlorophenol	EPA 525.2	<i>Subcontracted</i>	60
Prometon	EPA 525.2	<i>Subcontracted</i>	60
Propachlor	EPA 525.2	<i>Subcontracted</i>	60
Simazine	EPA 525.2	<i>Subcontracted</i>	60
Terbacil	EPA 525.2	<i>Subcontracted</i>	60
Tetrachlorobiphenyl	EPA 525.2	<i>Subcontracted</i>	60
trans-Nonachlor	EPA 525.2	<i>Subcontracted</i>	60
Trichlorobiphenyl	EPA 525.2	<i>Subcontracted</i>	60
Trifluralin	EPA 525.2	<i>Subcontracted</i>	60
Toxaphene	EPA 525.2	<i>Subcontracted</i>	60

General Index

Analyte	Method Number	Page #
Alkalinity	SM 2320B	13
Ammonia	SM 4500-NH ₃ H	14
Annual Inorganics and Metals (18 parameters)	Type 9 (Primary Inorganics and Metals Chemistry)	34
Arsenic (Total & Dissolved)	EPA 200.8	39
Biological Oxygen Demand (BOD)	<i>Subcontracted</i>	15
Bromate	EPA 300.1	16
Bromide	EPA 300.1	16,17
BTEX N	<i>Subcontracted</i>	55
Carbamates	<i>Subcontracted</i>	56
Chemistry Types (analytical groupings)	Total-Chemistry Type 2, Type 3, Type 9	36
Chlorate	EPA 300.1	16
Chloride	EPA 300.0, EPA 325.2	17
Chlorite	EPA 300.1	16
Chlorophyll-A	SM 10200H	48
Chromium-VI	<i>Subcontracted</i>	41
Color	SM 2120B	18
Conductivity	SM 2510B	19
Corrosivity	EPA 1110-CORR	20
Cyanide	EPA 335.4-CNCL	21
Fluoride	4500C-F	22
Geosmin and MIB	<i>Subcontracted</i>	57
Haloacetic Acids (HAA)	SM 6251B-HAA	50
Hardness	Calculation (EPA 200.7 – Mg ²⁺ & Ca ²⁺)	44
Herbicides	<i>Subcontracted</i>	58
Heterotrophic Plate Count	SM 9215B-HPC	66
Lead and Copper	EPA 200.8	42
<i>Legionella</i>	SM 9260J or IDEXX Legiolert	67
Mercury	EPA 245.1-T-HG, EPA 245.1-D-HG, EPA 7471B (T-HG or D-HG)	40
Metals, Total and Dissolved (analytical groupings)	Total-Metals Type 7, Type 9, RCRA Types, Filtered-Metals Type 3, Type 4, 6010-Metals (Soil)	47
New Drinking Water Source (46 parameters)	Type 7 (Total Inorganics and Metals Chemistry)	35
Nitrate and Nitrite	EPA 353.2-NO ₂ +NO ₃	23
Nitrite (only)	353.2-NO ₂	23
Nitrogen, Total Dissolved (DTN)	ASTM D 8083	24
Nutrients, Total and Dissolved (analytical groupings)	Total-Nutrients Type 2, Type 3, Type 4, Type 6; Filtered Nutrients Type 9	37
Odor (Threshold Odor Number, TON)	EPA 140.1	25
Oil and Grease	<i>Subcontracted</i>	59
Orthophosphate	EPA 300.1	
Perchlorate	<i>Subcontracted</i>	26
Periphyton	SM 10300C (modified)	53
Pesticides and SVOCs	<i>Subcontracted</i>	60

General Index

Analyte	Method Number	Page #
pH	EPA 150.1	XX
Phosphate	EPA 365.1	XX
Polychlorinated Biphenyls and Organochlorine Pesticides	<i>Subcontracted</i>	XX
Silica	SM 4500-SiO ₂ F	XX
Sulfate	EPA 300.0 or EPA 375.2	XX
Sulfide	EPA 376.2	XX
Surfactants	<i>Subcontracted</i>	XX
Total Coliforms and <i>E. coli</i>	SM 9223B	XX
Total Dissolved Solids (TDS)	SM 2540C	XX
Total Organic Carbon (TOC)	SM 5310B	XX
Total Petroleum Hydrocarbons (TPH)	<i>Subcontracted</i>	XX
Total Suspended Solids (TSS)	SM 2540D	XX
Toxic Characteristic Leaching Procedure-Metals	TCLP-Metals	XX
Toxic Characteristic Leaching Procedure-Organics	<i>Subcontracted</i>	XX
Trihalomethanes (THM)	EPA 524.2-THM	XX
Turbidity	EPA 180.1	XX
UV-Absorbing Organic Constituents	<i>Subcontracted</i>	XX
VOCs	<i>Subcontracted</i>	XX
Volatile Suspended Solids (VSS)	SM 2540E	XX